

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031419\
 Data File : VX008099.D
 Acq On : 15 Mar 2019 04:40
 Operator : JC/SP
 Sample : K1829-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-MW16-001-M20W2-030419

Quant Time: Mar 15 05:48:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	285874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	434748	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	411784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	203852	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	145413	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
35) Dibromofluoromethane	5.51	113	140034	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
50) Toluene-d8	8.71	98	521597	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	192336	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
Target Compounds						
16) Acetone	2.45	43	5022	3.689	ug/l #	88
44) Trichloroethene	7.21	130	4834	1.529	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031419\
Data File : VX008099.D
Acq On : 15 Mar 2019 04:40
Operator : JC/SP
Sample : K1829-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-MW16-001-M20W2-030419

Quant Time: Mar 15 05:48:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

