

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027416.D
 Acq On : 14 Mar 2022 12:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Mar 15 06:01:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	329581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	509717	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	481134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	255706	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	178234	46.374	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.740%
35) Dibromofluoromethane	5.385	113	163218	49.625	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.260%
50) Toluene-d8	8.653	98	589027	47.438	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.880%
62) 4-Bromofluorobenzene	11.079	95	234001	49.446	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.900%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	156760	46.401	ug/l	100
3) Chloromethane	1.295	50	143440	40.377	ug/l	99
4) Vinyl Chloride	1.374	62	159989	43.818	ug/l	99
5) Bromomethane	1.599	94	67442	45.175	ug/l	99
6) Chloroethane	1.679	64	98321	46.253	ug/l	100
7) Trichlorofluoromethane	1.886	101	238238	46.855	ug/l	98
8) Diethyl Ether	2.136	74	99828	50.388	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.325	101	158283	47.616	ug/l	95
10) Methyl Iodide	2.453	142	205513	43.800	ug/l	99
11) Tert butyl alcohol	2.965	59	189939	252.088	ug/l	100
12) 1,1-Dichloroethene	2.319	96	148734	46.542	ug/l	93
13) Acrolein	2.233	56	142059	268.975	ug/l	98
14) Allyl chloride	2.660	41	235523	46.987	ug/l	96
15) Acrylonitrile	3.062	53	476561	240.332	ug/l	100
16) Acetone	2.380	43	381185	255.973	ug/l	98
17) Carbon Disulfide	2.508	76	364844	42.022	ug/l	100
18) Methyl Acetate	2.703	43	202938	48.033	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	540606	51.058	ug/l	99
20) Methylene Chloride	2.788	84	176178	46.921	ug/l	100
21) trans-1,2-Dichloroethene	3.093	96	162330	45.701	ug/l	97
22) Diisopropyl ether	3.764	45	490854	48.984	ug/l	94
23) Vinyl Acetate	3.721	43	2149246	256.381	ug/l	98
24) 1,1-Dichloroethane	3.611	63	290609	47.761	ug/l	99
25) 2-Butanone	4.556	43	628931	237.564	ug/l	99
26) 2,2-Dichloropropane	4.477	77	226851	51.998	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	200972	48.767	ug/l	93
28) Bromochloromethane	4.898	49	117666	46.478	ug/l	90
29) Tetrahydrofuran	5.001	42	414617	236.691	ug/l	95
30) Chloroform	5.093	83	314958	49.043	ug/l	100
31) Cyclohexane	5.471	56	244072	45.400	ug/l	96
32) 1,1,1-Trichloroethane	5.385	97	273544	49.836	ug/l	97
36) 1,1-Dichloropropene	5.690	75	217998	47.739	ug/l	97
37) Ethyl Acetate	4.715	43	237457	49.331	ug/l	100
38) Carbon Tetrachloride	5.678	117	243341	51.534	ug/l	97
39) Methylcyclohexane	7.379	83	282944	49.068	ug/l	99
40) Benzene	6.038	78	674024	48.462	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027416.D
 Acq On : 14 Mar 2022 12:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Mar 15 06:01:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	131204	50.555	ug/l	96
42) 1,2-Dichloroethane	6.086	62	238146	52.416	ug/l	100
43) Isopropyl Acetate	6.342	43	374566	51.620	ug/l	99
44) Trichloroethene	7.129	130	191910	50.238	ug/l	98
45) 1,2-Dichloropropane	7.434	63	175439	50.341	ug/l	100
46) Dibromomethane	7.580	93	130545	52.651	ug/l	96
47) Bromodichloromethane	7.824	83	246985	53.940	ug/l	99
48) Methyl methacrylate	7.696	41	187804	52.047	ug/l	94
49) 1,4-Dioxane	7.659	88	95453	1010.200	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	1254077	258.145	ug/l	98
52) Toluene	8.720	92	448417	50.574	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	269602	57.386	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	291800	55.331	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	196433	53.086	ug/l	99
56) Ethyl methacrylate	9.116	69	295970	54.227	ug/l	96
57) 1,3-Dichloropropane	9.311	76	309456	51.806	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	635354	230.561	ug/l	98
59) 2-Hexanone	9.433	43	971991	259.017	ug/l	95
60) Dibromochloromethane	9.525	129	215836	57.152	ug/l	96
61) 1,2-Dibromoethane	9.610	107	209372	54.792	ug/l	99
64) Tetrachloroethene	9.275	164	170600	50.597	ug/l	98
65) Chlorobenzene	10.080	112	510663	50.449	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	193952	54.804	ug/l	99
67) Ethyl Benzene	10.195	91	870044	49.841	ug/l	96
68) m/p-Xylenes	10.305	106	700552	102.313	ug/l	98
69) o-Xylene	10.640	106	347554	51.412	ug/l	98
70) Styrene	10.659	104	589842	52.565	ug/l	99
71) Bromoform	10.799	173	174032	60.214	ug/l	100
73) Isopropylbenzene	10.964	105	897105	47.970	ug/l	100
74) N-amyl acetate	10.842	43	332925	48.708	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.213	83	307310	47.819	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	345808	61.639	ug/l	87
77) Bromobenzene	11.201	156	237047	50.231	ug/l	93
78) n-propylbenzene	11.305	91	1021959	48.581	ug/l	97
79) 2-Chlorotoluene	11.366	91	624199	47.477	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	777652	48.502	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	95977	55.327	ug/l	98
82) 4-Chlorotoluene	11.457	91	718576	48.314	ug/l	97
83) tert-Butylbenzene	11.719	119	788656	49.210	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	783585	49.276	ug/l	100
85) sec-Butylbenzene	11.890	105	947496	49.019	ug/l	99
86) p-Isopropyltoluene	12.012	119	827690	50.577	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	443247	49.994	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	443890	49.450	ug/l	99
89) n-Butylbenzene	12.335	91	686743	50.100	ug/l	97
90) Hexachloroethane	12.542	117	134623	52.780	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	436158	49.895	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	69476	51.599	ug/l	97
93) 1,2,4-Trichlorobenzene	13.591	180	288117	50.212	ug/l	99
94) Hexachlorobutadiene	13.725	225	128854	51.213	ug/l	93
95) Naphthalene	13.780	128	966974	50.198	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	284927	49.690	ug/l	96


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027416.D
Acq On : 14 Mar 2022 12:49
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Mar 15 06:01:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 329581
Ion Ratio Lower Upper
168 100
99 53.7 41.7 62.5




#2
Dichlorodifluoromethane
Concen: 46.401 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 85 Resp: 156760
Ion Ratio Lower Upper
85 100
87 32.5 16.2 48.6



Abundance Scan 34 (1.294 min): VX027331.D\data.ms (-28)



#3

Chloromethane

Concen: 40.377 ug/l

RT: 1.295 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tgt Ion: 50 Resp: 143440

Ion Ratio Lower Upper


50 100

52 32.4 26.4 39.6


Abundance Scan 34 (1.295 min): VX027416.D\data.ms




Abundance



Abundance Scan 34 (1.295 min): VX027416.D\data.ms (-1)



Abundance Scan 47 (1.374 min): VX027331.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 43.818 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49


Tgt Ion: 62 Resp: 159989

Ion Ratio Lower Upper


62 100

64 32.4 25.5 38.3


Abundance Scan 47 (1.374 min): VX027416.D\data.ms






Abundance



Abundance Scan 47 (1.374 min): VX027416.D\data.ms (-1)







Abundance Scan 84 (1.599 min): VX027331.D\data.ms (-74)



#5
Bromomethane
Concen: 45.175 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 67442
Ion Ratio Lower Upper
94 100
96 93.5 75.3 112.9



#6
Chloroethane
Concen: 46.253 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 64 Resp: 98321
Ion Ratio Lower Upper
64 100
66 31.7 25.3 37.9









#7
 Trichlorofluoromethane
 Concen: 46.855 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. -0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 238238
 Ion Ratio Lower Upper
 101 100
 103 64.4 50.5 75.7



#8
 Diethyl Ether
 Concen: 50.388 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49

Tgt Ion: 74 Resp: 99828
 Ion Ratio Lower Upper
 74 100
 45 85.7 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.616 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 158283
Ion Ratio Lower Upper
101 100
85 42.2 33.5 50.3
151 84.1 61.8 92.6



#10
Methyl Iodide
Concen: 43.800 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:142 Resp: 205513
Ion Ratio Lower Upper
142 100
127 42.2 33.9 50.9
141 13.6 11.4 17.2





#11

Tert butyl alcohol

Concen: 252.088 ug/l

RT: 2.965 min Scan# 308

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 59 Resp: 189939

Ion Ratio Lower Upper

59 100

57 10.3 8.3 12.5



#12

1,1-Dichloroethene

Concen: 46.542 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 96 Resp: 148734


Ion Ratio Lower Upper

96 100

61 144.9 124.6 187.0

98 65.2 51.0 76.4









#13
Acrolein
Concen: 268.975 ug/l
RT: 2.233 min Scan# 189
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 142059
Ion Ratio Lower Upper
56 100
55 70.0 55.0 82.6



#14
Allyl chloride
Concen: 46.987 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 41 Resp: 235523
Ion Ratio Lower Upper
41 100
39 60.3 50.4 75.6
76 37.6 32.7 49.1









#15
Acrylonitrile
Concen: 240.332 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 476561
Ion Ratio Lower Upper
53 100
52 82.1 65.7 98.5
51 35.6 28.4 42.6



#16
Acetone
Concen: 255.973 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 43 Resp: 381185
Ion Ratio Lower Upper
43 100
58 34.6 28.8 43.2









#17
Carbon Disulfide
Concen: 42.022 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 364844
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8



#18
Methyl Acetate
Concen: 48.033 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 43 Resp: 202938
Ion Ratio Lower Upper
43 100
74 27.1 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 51.058 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 540606
Ion Ratio Lower Upper
73 100
57 21.2 16.8 25.2



#20
Methylene Chloride
Concen: 46.921 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 84 Resp: 176178
Ion Ratio Lower Upper
84 100
49 107.0 85.5 128.3
51 32.8 26.6 39.8
86 65.1 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 45.701 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 162330
Ion Ratio Lower Upper
96 100
61 129.6 105.4 158.2
98 65.5 50.0 75.0



#22
Diisopropyl ether
Concen: 48.984 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 45 Resp: 490854
Ion Ratio Lower Upper
45 100
43 70.5 50.8 76.2
87 33.7 26.6 39.8
59 12.8 10.5 15.7









#23
 Vinyl Acetate
 Concen: 256.381 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 2149246
 Ion Ratio Lower Upper
 43 100
 86 12.6 10.8 16.2



#24
 1,1-Dichloroethane
 Concen: 47.761 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49

Tgt Ion: 63 Resp: 290609
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.9 11.7
 100 4.8 2.4 7.2









#25
2-Butanone
Concen: 237.564 ug/l
RT: 4.556 min Scan# 50
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 628931
Ion Ratio Lower Upper
43 100
72 30.0 24.4 36.6



#26
2,2-Dichloropropane
Concen: 51.998 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 77 Resp: 226851
Ion Ratio Lower Upper
77 100
97 24.6 11.8 35.4









#27
cis-1,2-Dichloroethene
Concen: 48.767 ug/l
RT: 4.489 min Scan# 511
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 200972
Ion Ratio Lower Upper
96 100
61 129.6 0.0 283.8
98 64.0 0.0 129.8



#28
Bromochloromethane
Concen: 46.478 ug/l
RT: 4.898 min Scan# 625
Delta R.T. 0.001 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 49 Resp: 117666
Ion Ratio Lower Upper
49 100
129 2.5 0.0 4.4
130 98.1 70.7 106.1









#29
Tetrahydrofuran
Concen: 236.691 ug/l
RT: 5.001 min Scan# 64
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 42 Resp: 414617
Ion Ratio Lower Upper
42 100
72 51.1 44.1 66.1
71 47.7 40.5 60.7



#30
Chloroform
Concen: 49.043 ug/l
RT: 5.093 min Scan# 657
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 83 Resp: 314958
Ion Ratio Lower Upper
83 100
85 65.8 52.7 79.1









#31
Cyclohexane
Concen: 45.400 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 244072
Ion Ratio Lower Upper
56 100
69 33.6 26.7 40.1
84 91.1 77.2 115.8



#32
1,1,1-Trichloroethane
Concen: 49.836 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 97 Resp: 273544
Ion Ratio Lower Upper
97 100
99 64.0 51.8 77.6
61 38.8 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 46.374 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 178234

Ion Ratio Lower Upper

65 100

67 54.0 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 114 Resp: 509717


Ion Ratio Lower Upper

114 100

63 18.0 0.0 36.8

88 15.6 0.0 32.4









#35
Dibromofluoromethane
Concen: 49.625 ug/l
RT: 5.385 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 163218
Ion Ratio Lower Upper
113 100
111 101.4 82.8 124.2
192 20.9 16.9 25.3



#36
1,1-Dichloropropene
Concen: 47.739 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 75 Resp: 217998
Ion Ratio Lower Upper
75 100
110 39.0 18.3 54.8
77 31.6 24.8 37.2





#37

Ethyl Acetate

Concen: 49.331 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.006 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 237457

Ion Ratio Lower Upper

43 100

61 15.8 12.6 19.0

70 13.1 10.2 15.2



#38

Carbon Tetrachloride

Concen: 51.534 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 117 Resp: 243341


Ion Ratio Lower Upper

117 100

119 93.0 77.3 115.9

121 30.6 24.2 36.2





#39

Methylcyclohexane

Concen: 49.068 ug/l

RT: 7.379 min Scan# 10

Delta R.T. -0.006 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 282944

Ion Ratio Lower Upper

83 100

55 69.2 54.6 81.8

98 47.2 37.4 56.2



#40

Benzene

Concen: 48.462 ug/l

RT: 6.038 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX027416.D




Acq: 14 Mar 2022 12:49


Tgt Ion: 78 Resp: 674024

Ion Ratio Lower Upper

78 100

77 23.5 19.3 28.9









#41
Methacrylonitrile
Concen: 50.555 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 41 Resp: 131204
Ion Ratio Lower Upper
41 100
39 54.0 43.0 64.4
67 81.8 70.2 105.4
52 31.1 26.0 39.0



#42
1,2-Dichloroethane
Concen: 52.416 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 62 Resp: 238146
Ion Ratio Lower Upper
62 100
98 10.3 0.0 20.4









#43
Isopropyl Acetate
Concen: 51.620 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 374566
Ion Ratio Lower Upper
43 100
61 22.3 18.5 27.7
87 16.0 13.0 19.6



#44
Trichloroethene
Concen: 50.238 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:130 Resp: 191910
Ion Ratio Lower Upper
130 100
95 92.0 0.0 188.8





#45

1,2-Dichloropropane

Concen: 50.341 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 175439

Ion Ratio Lower Upper

63 100

65 32.1 25.8 38.6



#46

Dibromomethane

Concen: 52.651 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 93 Resp: 130545


Ion Ratio Lower Upper

93 100

95 83.7 65.8 98.6

174 112.2 85.4 128.2









#47
Bromodichloromethane
Concen: 53.940 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 246985
Ion Ratio Lower Upper
83 100
85 65.9 52.2 78.2
127 9.9 6.8 10.2



#48
Methyl methacrylate
Concen: 52.047 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 41 Resp: 187804
Ion Ratio Lower Upper
41 100
69 94.4 82.5 123.7
39 53.1 43.8 65.8





#49

1,4-Dioxane

Concen: 1010.200 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 95453

Ion Ratio Lower Upper

88 100

43 27.2 21.9 32.9

58 62.8 53.7 80.5



#50

Toluene-d8

Concen: 47.438 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX027416.D




Acq: 14 Mar 2022 12:49


Tgt Ion: 98 Resp: 589027

Ion Ratio Lower Upper

98 100

100 65.1 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 258.145 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 1254077
Ion Ratio Lower Upper
43 100
58 43.5 36.1 54.1



#52
Toluene
Concen: 50.574 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 92 Resp: 448417
Ion Ratio Lower Upper
92 100
91 170.6 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 57.386 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 75 Resp: 269602

Ion Ratio Lower Upper

75 100

77 31.1 25.4 38.0



#54

cis-1,3-Dichloropropene

Concen: 55.331 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 75 Resp: 291800

Ion Ratio Lower Upper

75 100

77 32.0 24.8 37.2

39 40.1 29.6 44.4









#55
1,1,2-Trichloroethane
Concen: 53.086 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 196433




Ion	Ratio	Lower	Upper
97	100		
83	83.6	66.2	99.4
85	54.4	42.9	64.3
99	62.1	49.0	73.6




#56
Ethyl methacrylate
Concen: 54.227 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 69 Resp: 295970

Ion	Ratio	Lower	Upper
69	100		
41	58.6	43.8	65.6
39	28.9	22.2	33.2









#57
1,3-Dichloropropane
Concen: 51.806 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 309456
Ion Ratio Lower Upper
76 100
78 32.4 25.8 38.8



#58
2-Chloroethyl Vinyl ether
Concen: 230.561 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 63 Resp: 635354
Ion Ratio Lower Upper
63 100
106 31.2 23.9 35.9





#59

2-Hexanone

Concen: 259.017 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 971991

Ion Ratio Lower Upper

43 100

58 60.3 31.9 95.7



#60

Dibromochloromethane

Concen: 57.152 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX027416.D




Acq: 14 Mar 2022 12:49


Tgt Ion: 129 Resp: 215836

Ion Ratio Lower Upper

129 100

127 77.3 40.3 120.8





#61

1,2-Dibromoethane

Concen: 54.792 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion:107 Resp: 209372

Ion Ratio Lower Upper

107 100

109 94.4 74.4 111.6



#62

4-Bromofluorobenzene

Concen: 49.446 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 95 Resp: 234001


Ion Ratio Lower Upper

95 100

174 84.7 0.0 163.0

176 81.9 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 481134
Ion Ratio Lower Upper
117 100
82 53.1 41.8 62.8
119 32.5 24.8 37.2



#64
Tetrachloroethene
Concen: 50.597 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:164 Resp: 170600
Ion Ratio Lower Upper
164 100
166 128.0 101.2 151.8
129 94.6 73.0 109.4
131 93.3 74.2 111.2









#65
Chlorobenzene
Concen: 50.449 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.001 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 510663
Ion Ratio Lower Upper
112 100
114 32.3 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 54.804 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:131 Resp: 193952
Ion Ratio Lower Upper
131 100
133 95.5 48.3 144.8
119 65.0 32.6 98.0









#67
Ethyl Benzene
Concen: 49.841 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 870044
Ion Ratio Lower Upper
91 100
106 33.0 24.6 36.8



#68
m/p-Xylenes
Concen: 102.313 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 106 Resp: 700552
Ion Ratio Lower Upper
106 100
91 196.9 160.4 240.6









#69
o-Xylene
Concen: 51.412 ug/l
RT: 10.640 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 347554
Ion Ratio Lower Upper
106 100
91 208.9 106.1 318.1



#70
Styrene
Concen: 52.565 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:104 Resp: 589842
Ion Ratio Lower Upper
104 100
78 47.6 38.6 57.8
103 54.6 43.8 65.8









#71
Bromoform
Concen: 60.214 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 174032
Ion Ratio Lower Upper
173 100
175 48.1 24.1 72.2
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:152 Resp: 255706
Ion Ratio Lower Upper
152 100
115 80.9 37.5 112.7
150 172.1 0.0 338.2









#73
Isopropylbenzene
Concen: 47.970 ug/l
RT: 10.964 min Scan# 1
Delta R.T. 0.001 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 897105
Ion Ratio Lower Upper
105 100
120 26.9 13.5 40.5



#74
N-ethyl acetate
Concen: 48.708 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 43 Resp: 332925
Ion Ratio Lower Upper
43 100
70 48.9 42.5 63.7
55 26.9 23.1 34.7
61 26.9 22.8 34.2









#75
1,1,2,2-Tetrachloroethane
Concen: 47.819 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 307310
Ion Ratio Lower Upper
83 100
131 11.2 5.1 15.3
85 65.0 32.6 97.7



#76
1,2,3-Trichloropropane
Concen: 61.639 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 75 Resp: 345808
Ion Ratio Lower Upper
75 100
77 31.0 19.4 58.2









#77
Bromobenzene
Concen: 50.231 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 237047
Ion Ratio Lower Upper
156 100
77 144.7 78.5 235.3
158 97.1 50.2 150.7



#78
n-propylbenzene
Concen: 48.581 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 91 Resp: 1021959
Ion Ratio Lower Upper
91 100
120 24.4 11.6 34.8









#79
2-Chlorotoluene
Concen: 47.477 ug/l
RT: 11.366 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 624199
Ion Ratio Lower Upper
91 100
126 35.4 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 48.502 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion: 105 Resp: 777652
Ion Ratio Lower Upper
105 100
120 49.3 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 55.327 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	75	Resp:	95977
Ion	Ratio	Lower	Upper
75	100		
53	89.3	69.0	103.4
89	46.3	37.0	55.6



#82
4-Chlorotoluene
Concen: 48.314 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:	91	Resp:	718576
Ion	Ratio	Lower	Upper
91	100		
126	31.5	14.9	44.5









#83
tert-Butylbenzene
Concen: 49.210 ug/l
RT: 11.719 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 788656
Ion Ratio Lower Upper
119 100
91 54.5 26.9 80.6
134 22.7 11.4 34.2



#84
1,2,4-Trimethylbenzene
Concen: 49.276 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:105 Resp: 783585
Ion Ratio Lower Upper
105 100
120 45.5 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 49.019 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 947496
 Ion Ratio Lower Upper
 105 100
 134 20.7 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 50.577 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49

Tgt Ion:119 Resp: 827690
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.0 39.0
 91 23.1 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 49.994 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 443247

Ion Ratio Lower Upper

146 100

111 39.9 19.7 59.1

148 63.9 31.6 94.7



#88

1,4-Dichlorobenzene

Concen: 49.450 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion:146 Resp: 443890


Ion Ratio Lower Upper

146 100

111 39.4 19.2 57.6

148 63.9 32.3 96.9





#89

n-Butylbenzene

Concen: 50.100 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 686743

Ion Ratio Lower Upper

91 100

92 53.6 27.8 83.3

134 27.7 13.3 39.8



#90

Hexachloroethane

Concen: 52.780 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX027416.D




Acq: 14 Mar 2022 12:49


Tgt Ion: 117 Resp: 134623

Ion Ratio Lower Upper

117 100

201 80.7 39.5 118.5





#91

1,2-Dichlorobenzene

Concen: 49.895 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 436158

Ion Ratio Lower Upper

146 100

111 41.5 21.0 63.0

148 63.7 32.6 97.7



#92

1,2-Dibromo-3-Chloropropane

Concen: 51.599 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. -0.000 min

Lab File: VX027416.D

Acq: 14 Mar 2022 12:49




Tgt Ion: 75 Resp: 69476


Ion Ratio Lower Upper

75 100

155 95.2 47.0 141.2

157 120.8 57.9 173.6









#93
 1,2,4-Trichlorobenzene
 Concen: 50.212 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. 0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:180 Resp: 288117
 Ion Ratio Lower Upper
 180 100
 182 94.7 47.7 143.1
 145 32.1 15.2 45.5



#94
 Hexachlorobutadiene
 Concen: 51.213 ug/l
 RT: 13.725 min Scan# 2073
 Delta R.T. 0.000 min
 Lab File: VX027416.D
 Acq: 14 Mar 2022 12:49

Tgt Ion:225 Resp: 128854
 Ion Ratio Lower Upper
 225 100
 223 61.9 34.2 102.5
 227 63.2 33.9 101.7









#95
Naphthalene
Concen: 50.198 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 966974
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.4
129 11.0 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 49.690 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX027416.D
Acq: 14 Mar 2022 12:49

Tgt Ion:180 Resp: 284927
Ion Ratio Lower Upper
180 100
182 95.2 50.0 149.8
145 33.7 16.2 48.5

