

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027418.D
 Acq On : 14 Mar 2022 13:48
 Operator : JC/MD
 Sample : VX0314WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0314WBL01

Quant Time: Mar 15 06:01:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	401783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	645993	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	598982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	269586	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	219233	46.791	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.580%
35) Dibromofluoromethane	5.391	113	207989	49.897	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.800%
50) Toluene-d8	8.653	98	764264	48.566	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.140%
62) 4-Bromofluorobenzene	11.085	95	272310	45.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027418.D
Acq On : 14 Mar 2022 13:48
Operator : JC/MD
Sample : VX0314WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0314WBL01

Quant Time: Mar 15 06:01:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

