

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027423.D
 Acq On : 14 Mar 2022 16:00
 Operator : JC/MD
 Sample : N1895-13DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP031022DL

Quant Time: Mar 15 06:03:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	320091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	525354	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	579209	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	242896	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206163	55.231	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 110.460%	
35) Dibromofluoromethane	5.391	113	175870	51.881	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.760%	
50) Toluene-d8	8.653	98	727156	56.819	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 113.640%	
62) 4-Bromofluorobenzene	11.079	95	273042	55.978	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 111.960%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.483	56	1659	0.318	ug/l #	66
39) Methylcyclohexane	7.385	83	3086	0.519	ug/l #	83
52) Toluene	8.720	92	206969	22.648	ug/l	99
67) Ethyl Benzene	10.195	91	1058662	50.377	ug/l	98
68) m/p-Xylenes	10.305	106	2001770	242.848	ug/l	95
69) o-Xylene	10.646	106	972864	119.543	ug/l	97
73) Isopropylbenzene	10.963	105	64135	3.610	ug/l	100
78) n-propylbenzene	11.305	91	100481	5.028	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	187129	12.287	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	729045	48.264	ug/l	100
95) Naphthalene	13.780	128	162054	8.856	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027423.D
Acq On : 14 Mar 2022 16:00
Operator : JC/MD
Sample : N1895-13DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP031022DL

Quant Time: Mar 15 06:03:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

