

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027427.D
 Acq On : 14 Mar 2022 17:34
 Operator : JC/MD
 Sample : N1895-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Mar 15 06:05:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	317670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	512551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	484155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	230223	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	204276	55.142	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.280%
35) Dibromofluoromethane	5.391	113	171430	51.834	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.660%
50) Toluene-d8	8.653	98	619531	49.618	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	231518	48.651	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027427.D
Acq On : 14 Mar 2022 17:34
Operator : JC/MD
Sample : N1895-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Mar 15 06:05:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

