

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027431.D
 Acq On : 14 Mar 2022 19:08
 Operator : JC/MD
 Sample : N1915-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 15 06:06:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	311731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	513385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	482275	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	227457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	205229	56.455	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 112.920%	
35) Dibromofluoromethane	5.391	113	173507	52.377	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.760%	
50) Toluene-d8	8.653	98	615696	49.231	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.460%	
62) 4-Bromofluorobenzene	11.085	95	228956	48.034	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.060%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027431.D
Acq On : 14 Mar 2022 19:08
Operator : JC/MD
Sample : N1915-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 15 06:06:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

