

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027435.D
 Acq On : 14 Mar 2022 20:43
 Operator : JC/MD
 Sample : N1895-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Mar 15 06:06:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	305912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	507682	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	476497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	219981	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	200199	56.119	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.240%
35) Dibromofluoromethane	5.391	113	170498	52.047	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.100%
50) Toluene-d8	8.653	98	607523	49.123	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.085	95	224976	47.729	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027435.D
Acq On : 14 Mar 2022 20:43
Operator : JC/MD
Sample : N1895-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Mar 15 06:06:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

