

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
 Data File : VX027437.D
 Acq On : 14 Mar 2022 21:30
 Operator : JC/MD
 Sample : N1924-09 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 33606

Quant Time: Mar 15 06:07:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	306112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	505878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	460492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	235615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	200906	56.280	ug/l	0.00
Spiked Amount 50.000	Range	61 - 141	Recovery	=	112.560%	
35) Dibromofluoromethane	5.391	113	169828	52.027	ug/l	0.00
Spiked Amount 50.000	Range	69 - 133	Recovery	=	104.060%	
50) Toluene-d8	8.653	98	600810	48.754	ug/l	0.00
Spiked Amount 50.000	Range	65 - 126	Recovery	=	97.500%	
62) 4-Bromofluorobenzene	11.085	95	214308	45.628	ug/l	0.00
Spiked Amount 50.000	Range	58 - 135	Recovery	=	91.260%	
Target Compounds						
						Qvalue
40) Benzene	6.044	78	30489	2.209	ug/l	96
52) Toluene	8.720	92	13050	1.483	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031422\
Data File : VX027437.D
Acq On : 14 Mar 2022 21:30
Operator : JC/MD
Sample : N1924-09 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
33606

Quant Time: Mar 15 06:07:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

