

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034455.D
 Acq On : 14 Mar 2023 11:24
 Operator : JC/MD
 Sample : 01879-05DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20230308DL

Quant Time: Mar 15 06:37:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	273825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	471261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	420820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	180629	45.343	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.680%
35) Dibromofluoromethane	5.385	113	147964	47.554	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.647	98	567601	48.866	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	260064	55.430	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.860%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1321	0.434	ug/l	96
44) Trichloroethene	7.123	130	211794	58.305	ug/l	100
64) Tetrachloroethene	9.275	164	1355	0.479	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034455.D
Acq On : 14 Mar 2023 11:24
Operator : JC/MD
Sample : 01879-05DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20230308DL

Quant Time: Mar 15 06:37:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

