

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034456.D
 Acq On : 14 Mar 2023 11:48
 Operator : JC/MD
 Sample : 01879-20DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20230309DL

Quant Time: Mar 15 06:37:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	263550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453831	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	406660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	203432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	175233	45.704	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.400%
35) Dibromofluoromethane	5.385	113	142727	47.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.260%
50) Toluene-d8	8.647	98	547494	48.946	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.079	95	211138	46.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.460%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	185906	63.459	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1376	0.477	ug/l #	70
27) cis-1,2-Dichloroethene	4.495	96	1041	0.282	ug/l	93
44) Trichloroethene	7.123	130	29204	8.348	ug/l	98
64) Tetrachloroethene	9.275	164	15586	5.706	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034456.D
Acq On : 14 Mar 2023 11:48
Operator : JC/MD
Sample : 01879-20DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20230309DL

Quant Time: Mar 15 06:37:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

