

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034457.D
 Acq On : 14 Mar 2023 12:11
 Operator : JC/MD
 Sample : 01879-18DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20230309DL

Quant Time: Mar 15 06:38:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	252337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	427996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	381793	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166582	45.378	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.760%
35) Dibromofluoromethane	5.391	113	133708	47.316	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.640%
50) Toluene-d8	8.647	98	514915	48.812	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	241706	56.725	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.460%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	214347	76.418	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1871	0.677	ug/l #	62
27) cis-1,2-Dichloroethene	4.489	96	1548	0.438	ug/l	90
44) Trichloroethene	7.129	130	60187	18.244	ug/l	96
64) Tetrachloroethene	9.275	164	43307	16.886	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034457.D
Acq On : 14 Mar 2023 12:11
Operator : JC/MD
Sample : 01879-18DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20230309DL

Quant Time: Mar 15 06:38:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

