

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034459.D
 Acq On : 14 Mar 2023 12:58
 Operator : JC/MD
 Sample : 01879-09DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20230308DL

Quant Time: Mar 15 06:39:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	257723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	440802	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	436223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	170046	45.353	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.700%		
35) Dibromofluoromethane	5.385	113	138998	47.759	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.520%		
50) Toluene-d8	8.647	98	532948	49.053	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.100%		
62) 4-Bromofluorobenzene	11.079	95	235407	53.642	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	107.280%		
Target Compounds						
20) Methylene Chloride	2.788	84	993	0.290	ug/l #	79
44) Trichloroethene	7.129	130	111826	32.912	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034459.D
Acq On : 14 Mar 2023 12:58
Operator : JC/MD
Sample : 01879-09DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20230308DL

Quant Time: Mar 15 06:39:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

