

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034463.D
 Acq On : 14 Mar 2023 14:33
 Operator : JC/MD
 Sample : 01879-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20230308

Quant Time: Mar 15 06:41:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	422248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	164600	45.919	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.840%
35) Dibromofluoromethane	5.385	113	133132	47.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.647	98	512672	49.261	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.079	95	196709	46.793	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.580%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	559	0.204	ug/l #	73
44) Trichloroethene	7.129	130	54097	16.621	ug/l	97
64) Tetrachloroethene	9.275	164	3542	1.377	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034463.D
Acq On : 14 Mar 2023 14:33
Operator : JC/MD
Sample : 01879-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20230308

Quant Time: Mar 15 06:41:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

