

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034469.D
 Acq On : 14 Mar 2023 16:54
 Operator : JC/MD
 Sample : 01892-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03

Quant Time: Mar 15 06:50:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	243904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	410577	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	360607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161563	45.532	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.060%
35) Dibromofluoromethane	5.385	113	128572	47.429	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.860%
50) Toluene-d8	8.647	98	493258	48.742	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.079	95	181143	44.316	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.640%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	766	0.242	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034469.D
Acq On : 14 Mar 2023 16:54
Operator : JC/MD
Sample : 01892-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03

Quant Time: Mar 15 06:50:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

