

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034470.D
 Acq On : 14 Mar 2023 17:17
 Operator : JC/MD
 Sample : 01890-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-031023

Quant Time: Mar 15 06:50:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	243361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	414839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	163309	46.127	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.260%		
35) Dibromofluoromethane	5.385	113	132818	48.492	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.980%		
50) Toluene-d8	8.647	98	504560	49.347	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.700%		
62) 4-Bromofluorobenzene	11.079	95	184800	44.746	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.500%		
Target Compounds						Qvalue
20) Methylene Chloride	2.788	84	920	0.284	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034470.D
Acq On : 14 Mar 2023 17:17
Operator : JC/MD
Sample : 01890-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-031023

Quant Time: Mar 15 06:50:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

