

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034472.D
 Acq On : 14 Mar 2023 18:04
 Operator : JC/MD
 Sample : 01880-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20230310

Quant Time: Mar 15 06:51:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	255182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	432918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165696	44.633	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.260%
35) Dibromofluoromethane	5.385	113	134302	46.986	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.980%
50) Toluene-d8	8.647	98	524419	49.147	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	193831	44.972	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	757	0.318	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	338988	119.507	ug/l	98
12) 1,1-Dichloroethene	2.319	96	3271	1.171	ug/l	89
27) cis-1,2-Dichloroethene	4.489	96	5904	1.653	ug/l	87
30) Chloroform	5.105	83	1533	0.269	ug/l	92
38) Carbon Tetrachloride	5.672	117	1188	0.272	ug/l	91
44) Trichloroethene	7.123	130	400200	119.930	ug/l	99
45) 1,2-Dichloropropane	7.446	63	799	0.238	ug/l #	69
64) Tetrachloroethene	9.275	164	32679	12.728	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034472.D
Acq On : 14 Mar 2023 18:04
Operator : JC/MD
Sample : 01880-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20230310

Quant Time: Mar 15 06:51:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

