

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034483.D
 Acq On : 15 Mar 2023 05:03
 Operator : JC/MD
 Sample : 01903-06 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20230310

Quant Time: Mar 15 07:00:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	251343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	434431	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	381376	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	169350	46.314	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.620%
35) Dibromofluoromethane	5.385	113	137785	48.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.646	98	523241	48.866	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	199090	46.032	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.060%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.330	101	2408	0.862	ug/l	94
12) 1,1-Dichloroethene	2.318	96	551	0.200	ug/l	84
27) cis-1,2-Dichloroethene	4.489	96	807	0.229	ug/l	82
44) Trichloroethene	7.122	130	193941	57.917	ug/l	95
64) Tetrachloroethene	9.274	164	1705	0.666	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034483.D
Acq On : 15 Mar 2023 05:03
Operator : JC/MD
Sample : 01903-06 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20230310

Quant Time: Mar 15 07:00:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

