

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034484.D
 Acq On : 15 Mar 2023 05:27
 Operator : JC/MD
 Sample : 01903-12 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20230312

Quant Time: Mar 15 07:00:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	247709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	419458	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	187628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165282	45.865	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.720%
35) Dibromofluoromethane	5.385	113	133448	48.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.647	98	512521	49.574	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.140%
62) 4-Bromofluorobenzene	11.079	95	198379	47.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.000%
Target Compounds						
29) Tetrahydrofuran	5.007	42	38664	25.797	ug/l	98
44) Trichloroethene	7.135	130	830	0.257	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034484.D
Acq On : 15 Mar 2023 05:27
Operator : JC/MD
Sample : 01903-12 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20230312

Quant Time: Mar 15 07:00:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

