

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034486.D
 Acq On : 15 Mar 2023 06:14
 Operator : JC/MD
 Sample : 01903-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-20230310

Quant Time: Mar 15 07:01:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	242916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	416304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	370743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	163239	46.192	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.380%
35) Dibromofluoromethane	5.385	113	131126	47.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.420%
50) Toluene-d8	8.647	98	505260	49.242	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.480%
62) 4-Bromofluorobenzene	11.079	95	224317	54.123	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.240%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	1526	0.917	ug/l #	87
20) Methylene Chloride	2.794	84	944	0.292	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034486.D
Acq On : 15 Mar 2023 06:14
Operator : JC/MD
Sample : 01903-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-20230310

Quant Time: Mar 15 07:01:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

