

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034487.D
 Acq On : 15 Mar 2023 06:37
 Operator : JC/MD
 Sample : 01903-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20230310

Quant Time: Mar 15 08:26:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	247470	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	422481	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173402	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166057	46.125	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.240%
35) Dibromofluoromethane	5.385	113	133573	47.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	8.647	98	515271	49.483	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.960%
62) 4-Bromofluorobenzene	11.079	95	196941	46.823	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034487.D
Acq On : 15 Mar 2023 06:37
Operator : JC/MD
Sample : 01903-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20230310

Quant Time: Mar 15 08:26:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

