

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034493.D
 Acq On : 15 Mar 2023 08:58
 Operator : JC/MD
 Sample : 01903-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20230312

Quant Time: Mar 15 14:22:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	419984	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	164019	45.729	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.460%
35) Dibromofluoromethane	5.385	113	131137	47.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.580%
50) Toluene-d8	8.647	98	501459	48.443	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.880%
62) 4-Bromofluorobenzene	11.079	95	188396	45.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.120%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1140320	416.090	ug/l	99
12) 1,1-Dichloroethene	2.325	96	7470	2.768	ug/l	79
27) cis-1,2-Dichloroethene	4.495	96	3328	0.964	ug/l	91
30) Chloroform	5.092	83	1158	0.211	ug/l	86
44) Trichloroethene	7.122	130	103824	32.072	ug/l	97
45) 1,2-Dichloropropane	7.433	63	2919	0.897	ug/l	97
64) Tetrachloroethene	9.275	164	59075	24.019	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034493.D
Acq On : 15 Mar 2023 08:58
Operator : JC/MD
Sample : 01903-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20230312

Quant Time: Mar 15 14:22:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

