

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034497.D
 Acq On : 15 Mar 2023 10:32
 Operator : JC/MD
 Sample : 01903-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D3-20230312

Quant Time: Mar 15 14:24:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	238377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	411643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	174442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161041	46.438	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.880%		
35) Dibromofluoromethane	5.385	113	127206	46.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.600%		
50) Toluene-d8	8.647	98	491422	48.435	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.880%		
62) 4-Bromofluorobenzene	11.079	95	227373	55.481	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	110.960%		
Target Compounds						
16) Acetone	2.380	43	1010	0.618	ug/l	91
44) Trichloroethene	7.129	130	2201	0.694	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034497.D
Acq On : 15 Mar 2023 10:32
Operator : JC/MD
Sample : 01903-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D3-20230312

Quant Time: Mar 15 14:24:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

