

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
 Data File : VX034500.D
 Acq On : 15 Mar 2023 11:42
 Operator : JC/MD
 Sample : 01903-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D2-20230312

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 03/20/2023
 Supervised By :Mahesh Dadoda 03/20/2023

Quant Time: Mar 15 14:25:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	238537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	408010	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	360361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161562	46.557	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.120%		
35) Dibromofluoromethane	5.385	113	128656	47.759	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.520%		
50) Toluene-d8	8.647	98	493260	49.049	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.100%		
62) 4-Bromofluorobenzene	11.079	95	184255	45.360	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.720%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.325	101	923	0.348	ug/l #	93
29) Tetrahydrofuran	5.013	42	19933	13.811	ug/l	99
44) Trichloroethene	7.123	130	826m	0.263	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031423\
Data File : VX034500.D
Acq On : 15 Mar 2023 11:42
Operator : JC/MD
Sample : 01903-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D2-20230312

Quant Time: Mar 15 14:25:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 03/20/2023
Supervised By :Mahesh Dadoda 03/20/2023

