

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031424\
 Data File : VX040636.D
 Acq On : 14 Mar 2024 12:23
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 15 05:42:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 13:21:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.891	128	16397	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	93247	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	85648	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	50646	30.926	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 103.100%	
60) 4-Bromofluorobenzene	11.079	95	43684	29.735	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 99.100%	
63) Toluene-d8	8.647	98	111913	29.567	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 98.567%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031424\
Data File : VX040636.D
Acq On : 14 Mar 2024 12:23
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 15 05:42:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 13:21:29 2024
Response via : Initial Calibration

