



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031424\
Data File : VX040629.D
Acq On : 14 Mar 2024 09:03
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Tue Mar 05 13:21:29 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	6991	PASS
75	95	30	60	58.6	18201	PASS
95	95	100	100	100.0	31077	PASS
96	95	5	9	6.8	2106	PASS
173	174	0.00	2	0.9	210	PASS
174	95	50	100	71.8	22323	PASS
175	174	5	9	7.9	1767	PASS
176	174	95	101	98.1	21909	PASS
177	176	5	9	6.7	1474	PASS