

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031519\  
 Data File : VX008120.D  
 Acq On : 15 Mar 2019 21:43  
 Operator : JC/SP  
 Sample : K1856-05  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-HN-MW29D-20190310

Quant Time: Mar 16 01:55:47 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X031419W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Mar 15 03:56:35 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 248635   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.86  | 114  | 376721   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 357124   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 178728   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 125466   | 50.02 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 100.04% |          |
| 35) Dibromofluoromethane    | 5.51  | 113  | 122130   | 51.12 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 102.24% |          |
| 50) Toluene-d8              | 8.71  | 98   | 451643   | 49.71 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 99.42%  |          |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 167312   | 51.40 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 102.80% |          |
| Target Compounds            |       |      |          |       |         |          |
| 16) Acetone                 | 2.45  | 43   | 3083     | 2.604 | ug/l    | 92       |
| 44) Trichloroethene         | 7.21  | 130  | 3550     | 1.296 | ug/l    | 77       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031519\  
Data File : VX008120.D  
Acq On : 15 Mar 2019 21:43  
Operator : JC/SP  
Sample : K1856-05  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-HN-MW29D-20190310

Quant Time: Mar 16 01:55:47 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X031419W.M  
Quant Title : SW846 8260  
QLast Update : Fri Mar 15 03:56:35 2019  
Response via : Initial Calibration

