

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031519\
 Data File : VX008124.D
 Acq On : 15 Mar 2019 23:30
 Operator : JC/SP
 Sample : K1856-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-AOC22-MW09-20190311

Quant Time: Mar 16 02:03:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	244703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	372362	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	352776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176997	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	123480	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
35) Dibromofluoromethane	5.51	113	119784	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	8.71	98	447267	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	11.13	95	162898	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
Target Compounds						
16) Acetone	2.45	43	1716	1.473	ug/l	95
44) Trichloroethene	7.21	130	4291	1.584	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031519\
Data File : VX008124.D
Acq On : 15 Mar 2019 23:30
Operator : JC/SP
Sample : K1856-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-AOC22-MW09-20190311

Quant Time: Mar 16 02:03:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

