

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031519\
 Data File : VX008127.D
 Acq On : 16 Mar 2019 00:51
 Operator : JC/SP
 Sample : K1856-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW302D-20190311

Quant Time: Mar 16 02:21:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	388099	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	369324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	181592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	129617	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
35) Dibromofluoromethane	5.51	113	125838	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	8.71	98	465149	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
62) 4-Bromofluorobenzene	11.13	95	166564	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
Target Compounds						
12) 1,1-Dichloroethene	2.38	96	1750	0.894	ug/l	90
16) Acetone	2.45	43	2561	2.096	ug/l	97
44) Trichloroethene	7.21	130	15871	5.622	ug/l	96
64) Tetrachloroethene	9.33	164	1157	0.387	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031519\
Data File : VX008127.D
Acq On : 16 Mar 2019 00:51
Operator : JC/SP
Sample : K1856-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW302D-20190311

Quant Time: Mar 16 02:21:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

