

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031519\
 Data File : VX008108.D
 Acq On : 15 Mar 2019 15:23
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Mar 16 01:05:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X030819W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 08 12:01:36 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	92	0.00
2 M	Dichlorodifluoromethane	1.781	1.598	10.3	82	0.00
3 M	Chloromethane	1.980	1.469	25.8	69	0.00
4 M	Vinyl Chloride	1.962	1.550	21.0	77	0.00
5 M	Bromomethane	0.978	1.037	-6.0	91	0.00
6 M	Chloroethane	1.146	0.976	14.8	81	0.00
7 M	Trichlorofluoromethane	2.443	2.137	12.5	86	0.00
8 T	Diethyl Ether	1.258	1.144	9.1	84	0.00
9	1,1,2-Trichlorotrifluoroeth	1.837	1.725	6.1	88	0.00
10 M	1,1-Dichloroethene	1.741	1.641	5.7	87	0.00
11	Methyl Iodide	2.725	2.731	-0.2	93	0.00
12	Methyl Acetate	2.753	2.243	18.5	78	0.00
13 M	Acrolein	0.323	0.271	16.1	84	0.00
14 M	Acrylonitrile	1.169	1.021	12.7	79	0.00
15 M	Acetone	0.422	0.296	29.9	56	0.00
16 M	Carbon Disulfide	5.462	4.682	14.3	78	0.00
17	Allyl chloride	3.569	2.772	22.3	72	0.00
18 M	Methylene Chloride	1.983	1.789	9.8	83	0.00
19 M	trans-1,2-Dichloroethene	1.842	1.740	5.5	87	0.00
20 T	Diisopropyl ether	6.529	5.699	12.7	79	0.00
21 M	1,1-Dichloroethane	3.551	3.257	8.3	84	0.00
22 M	cis-1,2-Dichloroethene	2.125	2.105	0.9	92	0.00
23 M	tert-Butyl Alcohol	0.471	0.455	3.4	85	0.01
24 M	Methyl tert-Butyl Ether	5.985	5.484	8.4	85	0.00
25 M	Chloroform	3.352	3.245	3.2	89	0.00
26	Cyclohexane	3.281	2.855	13.0	81	0.00
27 s	1,2-Dichloroethane-d4	1.972	1.844	6.5	86	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
29	1,1-Dichloropropene	0.476	0.448	5.9	86	0.00
30 M	2-Butanone	0.321	0.280	12.8	73	0.00
31	2,2-Dichloropropane	0.439	0.417	5.0	90	0.00
32 M	1,1,1-Trichloroethane	0.514	0.521	-1.4	93	0.00
33 M	Carbon Tetrachloride	0.456	0.456	0.0	93	0.00
34 M	Benzene	1.458	1.409	3.4	88	0.00
35	Methacrylonitrile	0.313	0.277	11.5	80	0.00
36 M	1,2-Dichloroethane	0.474	0.456	3.8	87	0.00
37 M	Trichloroethene	0.395	0.402	-1.8	94	0.00
38	Methylcyclohexane	0.605	0.554	8.4	86	0.00
39 M	1,2-Dichloropropane	0.380	0.362	4.7	87	0.00
40	Dibromomethane	0.241	0.236	2.1	90	0.00
41 M	Bromodichloromethane	0.456	0.427	6.4	87	0.00
42 M	Vinyl Acetate	1.039	0.928	10.7	79	0.00
43	Ethyl Acetate	0.559	0.509	8.9	82	0.00
44	Isopropyl Acetate	0.853	0.763	10.6	82	0.00
45 T	1,4-Dioxane	0.009	0.010#	-11.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031519\
 Data File : VX008108.D
 Acq On : 15 Mar 2019 15:23
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Mar 16 01:05:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X030819W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 08 12:01:36 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.448	0.401	10.5	81	0.00
47	n-amyl Acetate	0.735	0.637	13.3	78	0.00
48 M	t-1,3-Dichloropropene	0.494	0.466	5.7	90	0.00
49 T	cis-1,3-Dichloropropene	0.559	0.529	5.4	89	0.00
50 M	1,1,2-Trichloroethane	0.368	0.369	-0.3	90	0.00
51	Ethyl methacrylate	0.557	0.535	3.9	87	0.00
52	1,3-Dichloropropane	0.607	0.594	2.1	88	0.00
53 M	Dibromochloromethane	0.378	0.379	-0.3	93	0.00
54 M	1,2-Dibromoethane	0.384	0.392	-2.1	91	0.00
55 M	2-Chloroethyl vinyl ether	0.259	0.235	9.3	85	0.00
56 M	Bromoform	0.302	0.304	-0.7	95	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
58 M	4-Methyl-2-Pentanone	0.618	0.587	5.0	81	0.00
59 M	2-Hexanone	0.488	0.447	8.4	76	0.00
60 S	4-Bromofluorobenzene	0.479	0.465	2.9	89	0.00
61 M	Tetrachloroethene	0.438	0.471	-7.5	96	0.00
62 M	Toluene	1.752	1.729	1.3	89	0.00
63 S	Toluene-d8	1.373	1.342	2.3	90	0.00
64 M	Chlorobenzene	1.116	1.127	-1.0	90	0.00
65	1,1,1,2-Tetrachloroethane	0.400	0.417	-4.2	96	0.00
66 M	Ethyl Benzene	1.882	1.801	4.3	87	0.00
67 M	m/p-Xylenes	0.721	0.709	1.7	89	0.00
68 M	o-Xylene	0.701	0.696	0.7	90	0.00
69 M	Styrene	1.154	1.153	0.1	90	0.00
70	Isopropylbenzene	1.872	1.856	0.9	89	0.00
71 M	1,1,2,2-Tetrachloroethane	0.618	0.623	-0.8	88	0.00
72	1,2,3-Trichloropropane	0.583	0.601	-3.1	90	0.00
73	Bromobenzene	0.498	0.516	-3.6	93	0.00
74	n-propylbenzene	2.112	2.015	4.6	86	0.00
75	2-Chlorotoluene	1.266	1.224	3.3	87	0.00
76	1,3,5-Trimethylbenzene	1.551	1.537	0.9	89	0.00
77	t-1,4-Dichloro-2-butene	0.180	0.173	3.9	92	0.00
78	4-Chlorotoluene	1.431	1.375	3.9	86	0.00
79	tert-butylbenzene	1.665	1.637	1.7	89	0.00
80	1,2,4-Trimethylbenzene	1.575	1.569	0.4	90	0.00
81	sec-Butylbenzene	1.865	1.803	3.3	88	0.00
82	p-Isopropyltoluene	1.665	1.637	1.7	89	0.00
83 M	1,3-Dichlorobenzene	0.886	0.898	-1.4	93	0.00
84 M	1,4-Dichlorobenzene	0.887	0.888	-0.1	91	0.00
85	n-Butylbenzene	1.424	1.282	10.0	84	0.00
86 T	Hexachloroethane	0.271	0.258	4.8	91	0.00
87 M	1,2-Dichlorobenzene	0.876	0.907	-3.5	94	0.00
88	1,2-Dibromo-3-Chloropropane	0.126	0.126	0.0	90	0.00
89	1,2,4-Trichlorobenzene	0.629	0.624	0.8	91	0.00
90	Hexachlorobutadiene	0.347	0.345	0.6	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031519\
Data File : VX008108.D
Acq On : 15 Mar 2019 15:23
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Mar 16 01:05:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X030819W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 08 12:01:36 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.804	1.850	-2.5	93	0.00
92	1,2,3-Trichlorobenzene	0.626	0.637	-1.8	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0