


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031519\
Data File : VX008109.D
Acq On : 15 Mar 2019 15:51
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Title : SW846 8260
Last Update : Fri Mar 15 03:56:35 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.4	10029	PASS
75	95	30	60	44.3	28784	PASS
95	95	100	100	100.0	64946	PASS
96	95	5	9	6.8	4432	PASS
173	174	0.00	2	1.4	825	PASS
174	95	50	100	91.9	59656	PASS
175	174	5	9	7.8	4665	PASS
176	174	95	101	96.7	57696	PASS
177	176	5	9	6.7	3869	PASS