

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021194.D
 Acq On : 15 Mar 2021 12:13
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 16 02:59:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	91442	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	152842	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	134132	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	65291	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	75510	48.43	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.86%
35) Dibromofluoromethane	5.470	113	52724	50.14	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.28%
50) Toluene-d8	8.694	98	187428	49.62	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.24%
62) 4-Bromofluorobenzene	11.118	95	72483	50.35	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.70%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.187	85	57678	49.15	ug/l 100
3) Chloromethane	1.317	50	54441	45.81	ug/l 98
4) Vinyl Chloride	1.393	62	56453	48.35	ug/l 99
5) Bromomethane	1.623	94	32749	48.68	ug/l 99
6) Chloroethane	1.705	64	33517	48.38	ug/l 100
7) Trichlorofluoromethane	1.911	101	88955	48.99	ug/l 97
8) Diethyl Ether	2.170	74	30097	48.83	ug/l 80
9) 1,1,2-Trichlorotrifluo...	2.364	101	50065	50.52	ug/l 96
10) Methyl Iodide	2.493	142	60561	45.59	ug/l # 89
11) Tert butyl alcohol	3.023	59	61785	227.48	ug/l 97
12) 1,1-Dichloroethene	2.352	96	47517	48.29	ug/l 86
13) Acrolein	2.276	56	36681	219.77	ug/l 98
14) Allyl chloride	2.705	41	99494	46.31	ug/l # 87
15) Acrylonitrile	3.117	53	145227	242.47	ug/l 97
16) Acetone	2.423	43	161177	257.16	ug/l # 88
17) Carbon Disulfide	2.552	76	140148	48.14	ug/l 98
18) Methyl Acetate	2.752	43	70969	46.75	ug/l 90
19) Methyl tert-butyl Ether	3.170	73	183913	49.14	ug/l 97
20) Methylene Chloride	2.834	84	55160	46.62	ug/l # 86
21) trans-1,2-Dichloroethene	3.140	96	52861	48.90	ug/l 90
22) Diisopropyl ether	3.829	45	194897	49.62	ug/l 94
23) Vinyl Acetate	3.788	43	854169	248.75	ug/l # 92
24) 1,1-Dichloroethane	3.670	63	104045	49.15	ug/l 96
25) 2-Butanone	4.635	43	225605	250.20	ug/l # 89
26) 2,2-Dichloropropane	4.552	77	94563	49.12	ug/l 95
27) cis-1,2-Dichloroethene	4.564	96	60902	49.26	ug/l 88
28) Bromochloromethane	4.982	49	50712	47.98	ug/l 85
29) Tetrahydrofuran	5.094	42	138725	238.52	ug/l 86
30) Chloroform	5.176	83	107512	48.77	ug/l 99
31) Cyclohexane	5.547	56	93349	47.32	ug/l 91
32) 1,1,1-Trichloroethane	5.464	97	95257	48.81	ug/l 96
36) 1,1-Dichloropropene	5.770	75	79838	49.84	ug/l 96
37) Ethyl Acetate	4.794	43	86160	48.83	ug/l # 93
38) Carbon Tetrachloride	5.753	117	84675	48.14	ug/l 98
39) Methylcyclohexane	7.441	83	92209	51.26	ug/l 93
40) Benzene	6.111	78	221843	49.54	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021194.D
 Acq On : 15 Mar 2021 12:13
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 16 02:59:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.999	41	48476	47.66	ug/l	89
42) 1,2-Dichloroethane	6.158	62	94029	49.36	ug/l	94
43) Isopropyl Acetate	6.411	43	146347	49.45	ug/l #	91
44) Trichloroethene	7.188	130	59028	49.62	ug/l	96
45) 1,2-Dichloropropane	7.488	63	59579	49.23	ug/l	100
46) Dibromomethane	7.635	93	41817	49.60	ug/l	91
47) Bromodichloromethane	7.876	83	87684	49.61	ug/l	97
48) Methyl methacrylate	7.747	41	75938	49.87	ug/l #	82
49) 1,4-Dioxane	7.712	88	25163	979.33	ug/l #	88
51) 4-Methyl-2-Pentanone	8.618	43	423176	250.77	ug/l	88
52) Toluene	8.765	92	135683	50.25	ug/l	98
53) t-1,3-Dichloropropene	9.024	75	92192	50.35	ug/l	99
54) cis-1,3-Dichloropropene	8.418	75	99430	50.97	ug/l #	85
55) 1,1,2-Trichloroethane	9.194	97	55658	49.98	ug/l	98
56) Ethyl methacrylate	9.159	69	88349	51.09	ug/l #	75
57) 1,3-Dichloropropane	9.353	76	96465	50.07	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.288	63	225053	259.40	ug/l	95
59) 2-Hexanone	9.471	43	322267	253.72	ug/l	85
60) Dibromochloromethane	9.565	129	63934	50.26	ug/l	100
61) 1,2-Dibromoethane	9.653	107	58938	50.45	ug/l	100
64) Tetrachloroethene	9.318	164	55934	48.15	ug/l	91
65) Chlorobenzene	10.124	112	140290	50.43	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.206	131	55024	50.17	ug/l	99
67) Ethyl Benzene	10.235	91	263664	51.13	ug/l	96
68) m/p-Xylenes	10.341	106	189183	102.39	ug/l	92
69) o-Xylene	10.683	106	91890	51.23	ug/l	94
70) Styrene	10.694	104	156740	51.53	ug/l	95
71) Bromoform	10.841	173	41509	44.30	ug/l #	95
73) Isopropylbenzene	11.000	105	252122	49.74	ug/l	99
74) N-amyl acetate	10.883	43	122040	50.41	ug/l #	87
75) 1,1,2,2-Tetrachloroethane	11.253	83	77440	49.53	ug/l	99
76) 1,2,3-Trichloropropane	11.277	75	96128	59.38	ug/l	90
77) Bromobenzene	11.236	156	59906	49.69	ug/l	94
78) n-propylbenzene	11.341	91	298159	51.10	ug/l	97
79) 2-Chlorotoluene	11.406	91	175028	49.55	ug/l	99
80) 1,3,5-Trimethylbenzene	11.489	105	218254	51.30	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.059	75	27126	47.56	ug/l	88
82) 4-Chlorotoluene	11.494	91	211080	51.06	ug/l	95
83) tert-Butylbenzene	11.753	119	193050	49.78	ug/l	88
84) 1,2,4-Trimethylbenzene	11.789	105	217668	50.82	ug/l	96
85) sec-Butylbenzene	11.930	105	239689	50.77	ug/l	99
86) p-Isopropyltoluene	12.047	119	226044	52.19	ug/l	97
87) 1,3-Dichlorobenzene	12.006	146	109780	52.16	ug/l	98
88) 1,4-Dichlorobenzene	12.077	146	106405	49.18	ug/l	97
89) n-Butylbenzene	12.371	91	206642	52.03	ug/l	95
90) Hexachloroethane	12.577	117	40024	50.10	ug/l	89
91) 1,2-Dichlorobenzene	12.377	146	103878	49.99	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.983	75	19859	49.07	ug/l	77
93) 1,2,4-Trichlorobenzene	13.630	180	71018	53.40	ug/l	97
94) Hexachlorobutadiene	13.765	225	30281	53.02	ug/l	97
95) Naphthalene	13.818	128	224310	46.72	ug/l	100
96) 1,2,3-Trichlorobenzene	14.001	180	67486	47.01	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021194.D
Acq On : 15 Mar 2021 12:13
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Mar 16 02:59:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 16 02:59:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.629 min Scan# 772
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:168 Resp: 91442
Ion Ratio Lower Upper
168 100
99 67.2 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 49.15 ug/l
RT: 1.187 min Scan# 17
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 85 Resp: 57678
Ion Ratio Lower Upper
85 100
87 33.1 16.4 49.2









#3
Chloromethane
Concen: 45.81 ug/l
RT: 1.317 min Scan# 39
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 54441
Ion Ratio Lower Upper
50 100
52 31.9 26.4 39.6



#4
Vinyl Chloride
Concen: 48.35 ug/l
RT: 1.393 min Scan# 52
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 62 Resp: 56453
Ion Ratio Lower Upper
62 100
64 31.6 25.7 38.5









#5
Bromomethane
Concen: 48.68 ug/l
RT: 1.623 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 32749
Ion Ratio Lower Upper
94 100
96 94.6 76.4 114.6



#6
Chloroethane
Concen: 48.38 ug/l
RT: 1.705 min Scan# 105
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 64 Resp: 33517
Ion Ratio Lower Upper
64 100
66 31.6 25.4 38.0









#7
Trichlorofluoromethane
Concen: 48.99 ug/l
RT: 1.911 min Scan# 140
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 88955
Ion Ratio Lower Upper
101 100
103 64.1 53.2 79.8



#8
Diethyl Ether
Concen: 48.83 ug/l
RT: 2.170 min Scan# 184
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 74 Resp: 30097
Ion Ratio Lower Upper
74 100
45 107.8 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.52 ug/l
RT: 2.364 min Scan# 217
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 50065
Ion Ratio Lower Upper
101 100
85 46.7 34.9 52.3
151 73.1 61.4 92.2



#10
Methyl Iodide
Concen: 45.59 ug/l
RT: 2.493 min Scan# 239
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13


Tgt Ion:142 Resp: 60561
Ion Ratio Lower Upper
142 100
127 51.0 33.3 49.9#
141 14.3 11.4 17.0








#11
Tert butyl alcohol
Concen: 227.48 ug/l
RT: 3.023 min Scan# 329
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 59 Resp: 61785
Ion Ratio Lower Upper
59 100
57 10.9 7.8 11.8




#12
1,1-Dichloroethene
Concen: 48.29 ug/l
RT: 2.352 min Scan# 215
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13



Tgt Ion: 96 Resp: 47517
Ion Ratio Lower Upper
96 100
61 179.3 124.6 187.0
98 61.4 50.6 76.0









#13
Acrolein
Concen: 219.77 ug/l
RT: 2.276 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 36681
Ion Ratio Lower Upper
56 100
55 72.4 56.4 84.6



#14
Allyl chloride
Concen: 46.31 ug/l
RT: 2.705 min Scan# 275
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 41 Resp: 99494
Ion Ratio Lower Upper
41 100
39 74.3 51.6 77.4
76 30.3 30.5 45.7#









#15
Acrylonitrile
Concen: 242.47 ug/l
RT: 3.117 min Scan# 345
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 145227
Ion Ratio Lower Upper
53 100
52 84.0 65.3 97.9
51 37.7 28.2 42.2



#16
Acetone
Concen: 257.16 ug/l
RT: 2.423 min Scan# 227
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 43 Resp: 161177
Ion Ratio Lower Upper
43 100
58 27.7 27.8 41.8#









#17
Carbon Disulfide
Concen: 48.14 ug/l
RT: 2.552 min Scan# 249
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 140148
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8



#18
Methyl Acetate
Concen: 46.75 ug/l
RT: 2.752 min Scan# 283
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 43 Resp: 70969
Ion Ratio Lower Upper
43 100
74 22.1 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 49.14 ug/l
RT: 3.170 min Scan# 354
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 183913
Ion Ratio Lower Upper
73 100
57 22.7 17.1 25.7



#20
Methylene Chloride
Concen: 46.62 ug/l
RT: 2.834 min Scan# 297
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 84 Resp: 55160
Ion Ratio Lower Upper
84 100
49 133.3 88.6 132.8#
51 40.4 27.0 40.6
86 65.2 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 48.90 ug/l
RT: 3.140 min Scan# 349
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 96 Resp: 52861
Ion Ratio Lower Upper
96 100
61 150.9 107.7 161.5
98 64.9 50.4 75.6



#22
Diisopropyl ether
Concen: 49.62 ug/l
RT: 3.829 min Scan# 466
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 45 Resp: 194897
Ion Ratio Lower Upper
45 100
43 68.9 58.0 87.0
87 26.1 25.4 38.0
59 10.9 10.2 15.2









#23
 Vinyl Acetate
 Concen: 248.75 ug/l
 RT: 3.788 min Scan# 459
 Delta R.T. -0.000 min
 Lab File: VX021194.D
 Acq: 15 Mar 2021 12:13




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 854169
 Ion Ratio Lower Upper
 43 100
 86 9.1 9.9 14.9#



#24
 1,1-Dichloroethane
 Concen: 49.15 ug/l
 RT: 3.670 min Scan# 439
 Delta R.T. -0.000 min
 Lab File: VX021194.D
 Acq: 15 Mar 2021 12:13

Tgt Ion: 63 Resp: 104045
 Ion Ratio Lower Upper
 63 100
 98 6.1 3.8 11.2
 100 4.1 2.5 7.5





#25

2-Butanone

Concen: 250.20 ug/l

RT: 4.635 min Scan# 603

Delta R.T. -0.000 min

Lab File: VX021194.D





Acq: 15 Mar 2021 12:13

Tgt Ion: 43 Resp: 225605

Ion Ratio Lower Upper

43 100

72 23.6 23.7 35.5#



#26

2,2-Dichloropropane

Concen: 49.12 ug/l

RT: 4.552 min Scan# 589

Delta R.T. 0.006 min

Lab File: VX021194.D




Acq: 15 Mar 2021 12:13


Tgt Ion: 77 Resp: 94563

Ion Ratio Lower Upper

77 100

97 20.9 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 49.26 ug/l

RT: 4.564 min Scan# 591





Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13

Tgt Ion: 96 Resp: 60902

Ion	Ratio	Lower	Upper
96	100		
61	164.0	0.0	285.6
98	65.0	0.0	129.6



#28

Bromochloromethane

Concen: 47.98 ug/l

RT: 4.982 min Scan# 662




Delta R.T. -0.000 min


Lab File: VX021194.D

Acq: 15 Mar 2021 12:13

Tgt Ion: 49 Resp: 50712

Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	4.0
130	71.0	68.2	102.2





#29

Tetrahydrofuran

Concen: 238.52 ug/l

RT: 5.094 min Scan# 681

Delta R.T. 0.006 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13





Tgt Ion: 42 Resp: 138725

Ion Ratio Lower Upper

42 100

72 41.7 41.7 62.5

71 39.0 38.4 57.6



#30

Chloroform

Concen: 48.77 ug/l

RT: 5.176 min Scan# 695

Delta R.T. -0.000 min

Lab File: VX021194.D




Acq: 15 Mar 2021 12:13


Tgt Ion: 83 Resp: 107512

Ion Ratio Lower Upper

83 100


85 65.9 52.0 78.0








#31
Cyclohexane
Concen: 47.32 ug/l
RT: 5.547 min Scan# 758
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 93349
Ion Ratio Lower Upper
56 100
69 30.7 25.3 37.9
84 78.1 71.4 107.0




#32
1,1,1-Trichloroethane
Concen: 48.81 ug/l
RT: 5.464 min Scan# 744
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13



Tgt Ion: 97 Resp: 95257
Ion Ratio Lower Upper
97 100
99 64.2 51.5 77.3
61 49.7 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 48.43 ug/l

RT: 6.029 min Scan# 840

Delta R.T. -0.000 min

Lab File: VX021194.D





Acq: 15 Mar 2021 12:13

Tgt Ion: 65 Resp: 75510

Ion Ratio Lower Upper

65 100

67 50.2 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.829 min Scan# 976

Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13




Tgt Ion: 114 Resp: 152842


Ion Ratio Lower Upper

114 100

63 24.0 0.0 40.8

88 17.5 0.0 32.8









#35
Dibromofluoromethane
Concen: 50.14 ug/l
RT: 5.470 min Scan# 745
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:113 Resp: 52724
Ion Ratio Lower Upper
113 100
111 102.7 81.9 122.9
192 18.7 16.7 25.1



#36
1,1-Dichloropropene
Concen: 49.84 ug/l
RT: 5.770 min Scan# 796
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 75 Resp: 79838
Ion Ratio Lower Upper
75 100
110 33.1 18.1 54.4
77 30.6 25.3 37.9









#37
Ethyl Acetate
Concen: 48.83 ug/l
RT: 4.794 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 86160
Ion Ratio Lower Upper
43 100
61 13.3 12.8 19.2
70 9.9 10.5 15.7#



#38
Carbon Tetrachloride
Concen: 48.14 ug/l
RT: 5.753 min Scan# 793
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 117 Resp: 84675
Ion Ratio Lower Upper
117 100
119 93.0 76.2 114.2
121 31.1 24.8 37.2





#39

Methylcyclohexane

Concen: 51.26 ug/l

RT: 7.441 min Scan# 1080

Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13





Tgt Ion: 83 Resp: 92209

Ion Ratio Lower Upper

83 100

55 80.7 58.9 88.3

98 44.8 38.1 57.1



#40

Benzene

Concen: 49.54 ug/l

RT: 6.111 min Scan# 854

Delta R.T. -0.000 min

Lab File: VX021194.D




Acq: 15 Mar 2021 12:13


Tgt Ion: 78 Resp: 221843

Ion Ratio Lower Upper

78 100

77 24.3 18.9 28.3









#41
Methacrylonitrile
Concen: 47.66 ug/l
RT: 4.999 min Scan# 665
Delta R.T. -0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	41	Resp:	48476
Ion	Ratio	Lower	Upper
41	100		
39	58.1	42.5	63.7
67	66.4	65.0	97.6
52	30.5	25.3	37.9



#42
1,2-Dichloroethane
Concen: 49.36 ug/l
RT: 6.158 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:	62	Resp:	94029
Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	20.2









#43
Isopropyl Acetate
Concen: 49.45 ug/l
RT: 6.411 min Scan# 905
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 146347
Ion Ratio Lower Upper
43 100
61 18.3 18.2 27.4
87 12.0 12.5 18.7#



#44
Trichloroethene
Concen: 49.62 ug/l
RT: 7.188 min Scan# 1037
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:130 Resp: 59028
Ion Ratio Lower Upper
130 100
95 103.2 0.0 198.0





#45

1,2-Dichloropropane

Concen: 49.23 ug/l

RT: 7.488 min Scan# 1088

Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13

Tgt Ion: 63 Resp: 59579

Ion Ratio Lower Upper





63 100

65 31.9 25.7 38.5

Instrument :

MSVOA_X

ClientSampleId :



#46

Dibromomethane

Concen: 49.60 ug/l

RT: 7.635 min Scan# 1113

Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13




Tgt Ion: 93 Resp: 41817


Ion Ratio Lower Upper

93 100

95 81.5 66.8 100.2

174 89.3 83.1 124.7









#47
Bromodichloromethane
Concen: 49.61 ug/l
RT: 7.876 min Scan# 1154
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 87684
Ion Ratio Lower Upper
83 100
85 67.2 51.4 77.0
127 8.4 6.8 10.2



#48
Methyl methacrylate
Concen: 49.87 ug/l
RT: 7.747 min Scan# 1132
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 41 Resp: 75938
Ion Ratio Lower Upper
41 100
69 72.8 76.0 114.0#
39 63.6 44.6 67.0









#49
1,4-Dioxane
Concen: 979.33 ug/l
RT: 7.712 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 88 Resp: 25163
Ion Ratio Lower Upper
88 100
43 40.5 23.2 34.8#
58 75.0 55.0 82.4



#50
Toluene-d8
Concen: 49.62 ug/l
RT: 8.694 min Scan# 1293
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 98 Resp: 187428
Ion Ratio Lower Upper
98 100
100 64.8 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 250.77 ug/l
RT: 8.618 min Scan# 1280
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 423176
Ion Ratio Lower Upper
43 100
58 36.4 35.4 53.0



#52
Toluene
Concen: 50.25 ug/l
RT: 8.765 min Scan# 1305
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 92 Resp: 135683
Ion Ratio Lower Upper
92 100
91 174.2 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 50.35 ug/l
RT: 9.024 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 92192
Ion Ratio Lower Upper
75 100
77 31.8 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 50.97 ug/l
RT: 8.418 min Scan# 1246
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 75 Resp: 99430
Ion Ratio Lower Upper
75 100
77 31.0 25.7 38.5
39 59.4 35.1 52.7#





#55

1,1,2-Trichloroethane

Concen: 49.98 ug/l

RT: 9.194 min Scan# 1378

Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13

Tgt Ion: 97 Resp: 55658





Ion Ratio Lower Upper

97 100

83 88.6 69.3 103.9

85 56.7 44.3 66.5

99 60.3 50.3 75.5



#56

Ethyl methacrylate

Concen: 51.09 ug/l

RT: 9.159 min Scan# 1372

Delta R.T. -0.000 min

Lab File: VX021194.D

Acq: 15 Mar 2021 12:13




Tgt Ion: 69 Resp: 88349


Ion Ratio Lower Upper

69 100

41 75.0 46.8 70.2#

39 47.1 24.7 37.1#









#57
1,3-Dichloropropane
Concen: 50.07 ug/l
RT: 9.353 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 96465
Ion Ratio Lower Upper
76 100
78 32.8 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 259.40 ug/l
RT: 8.288 min Scan# 1224
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 63 Resp: 225053
Ion Ratio Lower Upper
63 100
106 26.8 23.5 35.3









#59
2-Hexanone
Concen: 253.72 ug/l
RT: 9.471 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 322267
Ion Ratio Lower Upper
43 100
58 50.7 30.9 92.8



#60
Dibromochloromethane
Concen: 50.26 ug/l
RT: 9.565 min Scan# 1441
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 129 Resp: 63934
Ion Ratio Lower Upper
129 100
127 77.2 38.4 115.2









#61
1,2-Dibromoethane
Concen: 50.45 ug/l
RT: 9.653 min Scan# 1456
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 58938
Ion Ratio Lower Upper
107 100
109 94.8 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 50.35 ug/l
RT: 11.118 min Scan# 1705
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 95 Resp: 72483
Ion Ratio Lower Upper
95 100
174 70.7 0.0 154.6
176 69.0 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.094 min Scan# 1531
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 134132
Ion Ratio Lower Upper
117 100
82 62.7 47.4 71.2
119 32.4 26.6 39.8



#64
Tetrachloroethene
Concen: 48.15 ug/l
RT: 9.318 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:164 Resp: 55934
Ion Ratio Lower Upper
164 100
166 124.6 103.4 155.2
129 102.6 70.8 106.2
131 99.9 72.5 108.7









#65
Chlorobenzene
Concen: 50.43 ug/l
RT: 10.124 min Scan# 1536
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 140290
Ion Ratio Lower Upper
112 100
114 30.6 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 50.17 ug/l
RT: 10.206 min Scan# 1550
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:131 Resp: 55024
Ion Ratio Lower Upper
131 100
133 96.7 47.7 143.1
119 64.3 32.1 96.5









#67
Ethyl Benzene
Concen: 51.13 ug/l
RT: 10.235 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 263664
Ion Ratio Lower Upper
91 100
106 28.5 24.6 37.0



#68
m/p-Xylenes
Concen: 102.39 ug/l
RT: 10.341 min Scan# 1573
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:106 Resp: 189183
Ion Ratio Lower Upper
106 100
91 215.9 162.8 244.2









#69
o-Xylene
Concen: 51.23 ug/l
RT: 10.683 min Scan# 1631
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 91890
Ion Ratio Lower Upper
106 100
91 225.2 108.1 324.2



#70
Styrene
Concen: 51.53 ug/l
RT: 10.694 min Scan# 1633
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:104 Resp: 156740
Ion Ratio Lower Upper
104 100
78 56.3 40.0 60.0
103 54.7 43.4 65.0









#71
Bromoform
Concen: 44.30 ug/l
RT: 10.841 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 41509
Ion Ratio Lower Upper
173 100
175 51.3 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.065 min Scan# 1866
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:152 Resp: 65291
Ion Ratio Lower Upper
152 100
115 88.7 41.1 123.3
150 169.9 0.0 349.0









#73
Isopropylbenzene
Concen: 49.74 ug/l
RT: 11.000 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 252122
Ion Ratio Lower Upper
105 100
120 25.6 13.2 39.6



#74
N-ethyl acetate
Concen: 50.41 ug/l
RT: 10.883 min Scan# 1665
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 43 Resp: 122040
Ion Ratio Lower Upper
43 100
70 36.8 39.0 58.6#
55 22.9 21.5 32.3
61 22.1 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 49.53 ug/l
RT: 11.253 min Scan# 1728
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 77440
Ion Ratio Lower Upper
83 100
131 10.3 5.1 15.2
85 64.2 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 59.38 ug/l
RT: 11.277 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 75 Resp: 96128
Ion Ratio Lower Upper
75 100
77 32.4 19.1 57.5









#77
Bromobenzene
Concen: 49.69 ug/l
RT: 11.236 min Scan# 1725
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
156	100		
77	177.6	82.8	248.6
158	97.1	49.1	147.4



#78
n-propylbenzene
Concen: 51.10 ug/l
RT: 11.341 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion	Ratio	Lower	Upper
91	100		
120	21.8	11.5	34.5









#79
2-Chlorotoluene
Concen: 49.55 ug/l
RT: 11.406 min Scan# 1754
Delta R.T. 0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 175028
Ion Ratio Lower Upper
91 100
126 32.4 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 51.30 ug/l
RT: 11.489 min Scan# 1768
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13


Tgt Ion: 105 Resp: 218254
Ion Ratio Lower Upper
105 100
120 46.2 23.8 71.5








#81
trans-1,4-Dichloro-2-butene
Concen: 47.56 ug/l
RT: 11.059 min Scan# 1695
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 75 Resp: 27126
Ion Ratio Lower Upper
75 100
53 108.5 74.5 111.7
89 44.9 37.9 56.9




#82
4-Chlorotoluene
Concen: 51.06 ug/l
RT: 11.494 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13



Tgt Ion: 91 Resp: 211080
Ion Ratio Lower Upper
91 100
126 27.2 14.9 44.5









#83
tert-Butylbenzene
Concen: 49.78 ug/l
RT: 11.753 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 193050
Ion Ratio Lower Upper
119 100
91 68.0 27.9 83.7
134 22.9 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 50.82 ug/l
RT: 11.789 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:105 Resp: 217668
Ion Ratio Lower Upper
105 100
120 42.6 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 50.77 ug/l
 RT: 11.930 min Scan# 1843
 Delta R.T. -0.000 min
 Lab File: VX021194.D
 Acq: 15 Mar 2021 12:13




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 239689
 Ion Ratio Lower Upper
 105 100
 134 19.4 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 52.19 ug/l
 RT: 12.047 min Scan# 1863
 Delta R.T. -0.000 min
 Lab File: VX021194.D
 Acq: 15 Mar 2021 12:13

Tgt Ion:119 Resp: 226044
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.9 38.7
 91 26.1 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 52.16 ug/l
RT: 12.006 min Scan# 1856
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Tgt Ion:146 Resp: 109780
Ion Ratio Lower Upper
146 100
111 42.1 20.2 60.5
148 62.5 31.9 95.7


Instrument :
MSVOA_X
ClientSampleId :



#88
1,4-Dichlorobenzene
Concen: 49.18 ug/l
RT: 12.077 min Scan# 1868
Delta R.T. -0.006 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:146 Resp: 106405
Ion Ratio Lower Upper
146 100
111 42.4 19.6 58.8
148 64.9 31.6 95.0









#89
n-Butylbenzene
Concen: 52.03 ug/l
RT: 12.371 min Scan# 1918
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 206642
Ion Ratio Lower Upper
91 100
92 51.0 27.4 82.0
134 24.0 13.2 39.5



#90
Hexachloroethane
Concen: 50.10 ug/l
RT: 12.577 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 117 Resp: 40024
Ion Ratio Lower Upper
117 100
201 70.8 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 49.99 ug/l
RT: 12.377 min Scan# 1919
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Tgt Ion:146 Resp: 103878
Ion Ratio Lower Upper
146 100
111 44.3 20.8 62.2
148 65.7 31.9 95.5


Instrument :
MSVOA_X
ClientSampled :



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.07 ug/l
RT: 12.983 min Scan# 2022
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion: 75 Resp: 19859
Ion Ratio Lower Upper
75 100
155 70.5 45.3 135.9
157 88.9 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 53.40 ug/l
RT: 13.630 min Scan# 2132
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:180 Resp: 71018
Ion Ratio Lower Upper
180 100
182 92.8 47.7 143.1
145 31.2 15.0 45.0



#94
Hexachlorobutadiene
Concen: 53.02 ug/l
RT: 13.765 min Scan# 2155
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:225 Resp: 30281
Ion Ratio Lower Upper
225 100
223 60.8 31.1 93.2
227 59.3 31.6 95.0









#95
Naphthalene
Concen: 46.72 ug/l
RT: 13.818 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:128 Resp: 224310
Ion Ratio Lower Upper
128 100
127 13.1 10.3 15.5
129 11.2 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 47.01 ug/l
RT: 14.001 min Scan# 2195
Delta R.T. -0.000 min
Lab File: VX021194.D
Acq: 15 Mar 2021 12:13

Tgt Ion:180 Resp: 67486
Ion Ratio Lower Upper
180 100
182 97.8 47.6 142.9
145 34.1 16.1 48.3

