

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021195.D
 Acq On : 15 Mar 2021 13:22
 Operator : JC/MD
 Sample : VX0315WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0315WBL01

Quant Time: Mar 16 02:59:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	75296	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	138906	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	136660	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	56232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	64771	50.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.90%	
35) Dibromofluoromethane	5.458	113	47987	50.21	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.42%	
50) Toluene-d8	8.694	98	179583	52.31	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.62%	
62) 4-Bromofluorobenzene	11.118	95	70512	53.90	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021195.D
Acq On : 15 Mar 2021 13:22
Operator : JC/MD
Sample : VX0315WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0315WBL01

Quant Time: Mar 16 02:59:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

