

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021198.D
 Acq On : 15 Mar 2021 14:47
 Operator : JC/MD
 Sample : M1583-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Mar 16 03:01:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	75078	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	136158	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	135760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	59148	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	62144	48.54	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.08%
35) Dibromofluoromethane	5.464	113	46488	49.62	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.24%
50) Toluene-d8	8.694	98	173215	51.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.94%
62) 4-Bromofluorobenzene	11.118	95	68911	53.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.48%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021198.D
Acq On : 15 Mar 2021 14:47
Operator : JC/MD
Sample : M1583-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Mar 16 03:01:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

