

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021203.D
 Acq On : 15 Mar 2021 16:36
 Operator : JC/MD
 Sample : M1596-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-318.9-20210305

Quant Time: Mar 16 03:03:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	74134	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.835	114	135878	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	130689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	57001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	64898	51.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.68%	
35) Dibromofluoromethane	5.476	113	47836	51.17	ug/l	0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.34%	
50) Toluene-d8	8.694	98	170644	50.81	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.62%	
62) 4-Bromofluorobenzene	11.118	95	67826	53.00	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 106.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021203.D
Acq On : 15 Mar 2021 16:36
Operator : JC/MD
Sample : M1596-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-318.9-20210305

Quant Time: Mar 16 03:03:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

