

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021207.D
 Acq On : 15 Mar 2021 18:07
 Operator : JC/MD
 Sample : M1677-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-3-12-2021

Quant Time: Mar 16 03:04:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	71971	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	133764	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	124098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51404	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	61484	50.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.20%
35) Dibromofluoromethane	5.470	113	44943	48.83	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.66%
50) Toluene-d8	8.694	98	165335	50.01	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.02%
62) 4-Bromofluorobenzene	11.118	95	61088	48.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.98%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021207.D
Acq On : 15 Mar 2021 18:07
Operator : JC/MD
Sample : M1677-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-3-12-2021

Quant Time: Mar 16 03:04:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

