

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021221.D
 Acq On : 16 Mar 2021 00:21
 Operator : JC/MD
 Sample : M1645-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 16 04:41:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	73274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	135080	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	133171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	57664	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	63069	50.48	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.96%
35) Dibromofluoromethane	5.470	113	45793	49.27	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.54%
50) Toluene-d8	8.694	98	172254	51.59	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.18%
62) 4-Bromofluorobenzene	11.118	95	67939	53.40	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021221.D
Acq On : 16 Mar 2021 00:21
Operator : JC/MD
Sample : M1645-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 16 04:41:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

