

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
 Data File : VX021231.D
 Acq On : 16 Mar 2021 04:15
 Operator : JC/MD
 Sample : M1666-02 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1858

Quant Time: Mar 16 04:45:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 15 11:55:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	70185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	134175	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	125697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	57336	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62674	52.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.74%
35) Dibromofluoromethane	5.464	113	45884	49.70	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.40%
50) Toluene-d8	8.694	98	165407	49.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.76%
62) 4-Bromofluorobenzene	11.118	95	66288	52.45	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.90%
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	1413	2.94	ug/l #	85
25) 2-Butanone	4.641	43	1927	2.78	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031521\
Data File : VX021231.D
Acq On : 16 Mar 2021 04:15
Operator : JC/MD
Sample : M1666-02 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1858

Quant Time: Mar 16 04:45:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031521W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 15 11:55:59 2021
Response via : Initial Calibration

