

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
 Data File : VX027442.D
 Acq On : 15 Mar 2022 12:57
 Operator : JC/MD
 Sample : VX0315MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0315MBL01

Quant Time: Mar 16 04:15:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	304935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	494179	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	458420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	211345	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	194165	54.602	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.200%
35) Dibromofluoromethane	5.385	113	165515	51.906	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.820%
50) Toluene-d8	8.647	98	589168	48.941	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.880%
62) 4-Bromofluorobenzene	11.079	95	215733	47.019	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
Data File : VX027442.D
Acq On : 15 Mar 2022 12:57
Operator : JC/MD
Sample : VX0315MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0315MBL01

Quant Time: Mar 16 04:15:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

