

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
 Data File : VX027456.D
 Acq On : 15 Mar 2022 18:42
 Operator : JC/MD
 Sample : N1950-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31339

Quant Time: Mar 16 04:21:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	377241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	622720	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	557382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	238798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	235151	53.453	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.900%			
35) Dibromofluoromethane	5.391	113	202999	50.520	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.040%			
50) Toluene-d8	8.653	98	732512	48.288	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.580%			
62) 4-Bromofluorobenzene	11.085	95	261559	45.240	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.480%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.959	59	27023	31.334	ug/l	95
16) Acetone	2.380	43	302590	176.628	ug/l	94
25) 2-Butanone	4.562	43	118154	38.991	ug/l	93
52) Toluene	8.720	92	11781	1.088	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
Data File : VX027456.D
Acq On : 15 Mar 2022 18:42
Operator : JC/MD
Sample : N1950-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31339

Quant Time: Mar 16 04:21:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

