

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
 Data File : VX027441.D
 Acq On : 15 Mar 2022 12:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 16 04:14:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.513	0.539	-5.1	83	0.00
3 P	Chloromethane	0.539	0.462	14.3	72	0.00
4 C	Vinyl Chloride	0.554	0.524	5.4#	77	0.00
5 T	Bromomethane	0.243	0.264	-8.6	92	0.00
6 T	Chloroethane	0.346	0.337	2.6	87	0.00
7 T	Trichlorofluoromethane	0.771	0.870	-12.8	93	0.00
8 T	Diethyl Ether	0.301	0.332	-10.3	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.575	-14.1	95	0.00
10 T	Methyl Iodide	0.712	0.747	-4.9	81	0.00
11 T	Tert butyl alcohol	0.114	0.121	-6.1	85	0.00
12 CM	1,1-Dichloroethene	0.485	0.538	-10.9#	89	0.00
13 T	Acrolein	0.080	0.094	-17.5	102	0.00
14 T	Allyl chloride	0.760	0.787	-3.6	83	0.00
15 T	Acrylonitrile	0.301	0.293	2.7	79	0.00
16 T	Acetone	0.256	0.275	-7.4	94	0.00
17 T	Carbon Disulfide	1.317	1.250	5.1	80	0.00
18 T	Methyl Acetate	0.641	0.652	-1.7	83	0.00
19 T	Methyl tert-butyl Ether	1.606	1.715	-6.8	86	0.00
20 T	Methylene Chloride	0.570	0.547	4.0	80	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.520	3.5	81	0.00
22 T	Diisopropyl ether	1.520	1.611	-6.0	84	0.00
23 T	Vinyl Acetate	1.272	1.410	-10.8	87	0.00
24 P	1,1-Dichloroethane	0.923	0.956	-3.6	84	0.00
25 T	2-Butanone	0.402	0.414	-3.0	84	0.00
26 T	2,2-Dichloropropane	0.662	0.790	-19.3	96	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.623	0.3	82	0.00
28 T	Bromochloromethane	0.384	0.382	0.5	81	0.00
29 T	Tetrahydrofuran	0.266	0.264	0.8	80	0.00
30 C	Chloroform	0.974	1.045	-7.3#	87	0.00
31 T	Cyclohexane	0.816	0.806	1.2	78	0.00
32 T	1,1,1-Trichloroethane	0.833	0.944	-13.3	90	0.00
33 S	1,2-Dichloroethane-d4	0.583	0.657	-12.7	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.323	0.348	-7.7	83	0.00
36 T	1,1-Dichloropropene	0.448	0.481	-7.4	84	0.00
37 T	Ethyl Acetate	0.472	0.503	-6.6	83	0.00
38 T	Carbon Tetrachloride	0.463	0.555	-19.9	92	0.00
39 T	Methylcyclohexane	0.566	0.592	-4.6	81	0.00
40 TM	Benzene	1.364	1.383	-1.4	80	0.00
41 T	Methacrylonitrile	0.255	0.277	-8.6	83	0.00
42 TM	1,2-Dichloroethane	0.446	0.547	-22.6	94	0.00
43 T	Isopropyl Acetate	0.712	0.798	-12.1	87	0.00
44 TM	Trichloroethene	0.375	0.399	-6.4	84	0.00
45 C	1,2-Dichloropropane	0.342	0.360	-5.3#	81	0.00
46 T	Dibromomethane	0.243	0.273	-12.3	87	0.00
47 T	Bromodichloromethane	0.449	0.531	-18.3	91	0.00
48 T	Methyl methacrylate	0.354	0.412	-16.4	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
 Data File : VX027441.D
 Acq On : 15 Mar 2022 12:19
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 16 04:14:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	79	0.00
50 S	Toluene-d8	1.218	1.273	-4.5	81	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.536	-12.4	85	0.00
52 CM	Toluene	0.870	0.929	-6.8#	82	0.00
53 T	t-1,3-Dichloropropene	0.461	0.555	-20.4	89	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.596	-15.3	86	0.00
55 T	1,1,2-Trichloroethane	0.363	0.395	-8.8	85	0.00
56 T	Ethyl methacrylate	0.535	0.597	-11.6	84	0.00
57 T	1,3-Dichloropropane	0.586	0.636	-8.5	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.270	0.296	-9.6	83	0.00
59 T	2-Hexanone	0.368	0.414	-12.5	84	0.00
60 T	Dibromochloromethane	0.370	0.447	-20.8	91	0.00
61 T	1,2-Dibromoethane	0.375	0.424	-13.1	86	0.00
62 S	4-Bromofluorobenzene	0.464	0.502	-8.2	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.350	0.387	-10.6	91	0.00
65 PM	Chlorobenzene	1.052	1.120	-6.5	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.427	-16.0	91	0.00
67 C	Ethyl Benzene	1.814	1.963	-8.2#	85	0.00
68 T	m/p-Xylenes	0.712	0.776	-9.0	86	0.00
69 T	o-Xylene	0.703	0.759	-8.0	85	0.00
70 T	Styrene	1.166	1.281	-9.9	84	0.00
71 P	Bromoform	0.300	0.356	-18.7	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.657	3.838	-4.9	86	0.00
74 T	N-amyl acetate	1.337	1.431	-7.0	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.257	1.233	1.9	83	0.00
76 T	1,2,3-Trichloropropane	1.097	1.147	-4.6	86	0.00
77 T	Bromobenzene	0.923	0.964	-4.4	87	0.00
78 T	n-propylbenzene	4.113	4.388	-6.7	86	0.00
79 T	2-Chlorotoluene	2.571	2.655	-3.3	86	0.00
80 T	1,3,5-Trimethylbenzene	3.135	3.298	-5.2	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.339	0.374	-10.3	87	0.00
82 T	4-Chlorotoluene	2.908	3.097	-6.5	87	0.00
83 T	tert-Butylbenzene	3.134	3.308	-5.6	86	0.00
84 T	1,2,4-Trimethylbenzene	3.109	3.304	-6.3	86	0.00
85 T	sec-Butylbenzene	3.780	4.062	-7.5	87	0.00
86 T	p-Isopropyltoluene	3.200	3.520	-10.0	88	0.00
87 T	1,3-Dichlorobenzene	1.734	1.801	-3.9	86	0.00
88 T	1,4-Dichlorobenzene	1.755	1.827	-4.1	88	0.00
89 T	n-Butylbenzene	2.680	3.010	-12.3	89	0.00
90 T	Hexachloroethane	0.499	0.570	-14.2	92	0.00
91 T	1,2-Dichlorobenzene	1.709	1.778	-4.0	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.285	-8.4	88	0.00
93 T	1,2,4-Trichlorobenzene	1.122	1.136	-1.2	84	0.00
94 T	Hexachlorobutadiene	0.492	0.519	-5.5	88	0.00
95 T	Naphthalene	3.767	3.796	-0.8	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031522\
Data File : VX027441.D
Acq On : 15 Mar 2022 12:19
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 16 04:14:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.121	1.141	-1.8	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6