

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
 Data File : VX034509.D
 Acq On : 15 Mar 2023 15:35
 Operator : JC/MD
 Sample : 01922-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-031423

Quant Time: Mar 16 05:20:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	234400	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	402062	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	350458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	158320	46.428	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.860%
35) Dibromofluoromethane	5.385	113	124723	46.984	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.960%
50) Toluene-d8	8.647	98	482592	48.698	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.079	95	176336	44.053	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.100%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	8339	5.191	ug/l	98
20) Methylene Chloride	2.794	84	3375	1.083	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
Data File : VX034509.D
Acq On : 15 Mar 2023 15:35
Operator : JC/MD
Sample : 01922-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-031423

Quant Time: Mar 16 05:20:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

