

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
 Data File : VX034512.D
 Acq On : 15 Mar 2023 16:45
 Operator : JC/MD
 Sample : 01922-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-05B-69-74-031423

Quant Time: Mar 16 05:21:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	242969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	422685	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	167396	47.358	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	94.720%	
35) Dibromofluoromethane	5.385	113	131672	47.181	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.360%	
50) Toluene-d8	8.647	98	508561	48.815	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	97.640%	
62) 4-Bromofluorobenzene	11.079	95	199776	47.474	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.940%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	4181	1.572	ug/l	99
20) Methylene Chloride	2.782	84	687	0.213	ug/l #	76
24) 1,1-Dichloroethane	3.611	63	7052	1.276	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	10687	3.143	ug/l	92
42) 1,2-Dichloroethane	6.104	62	4955	1.075	ug/l #	75
44) Trichloroethene	7.123	130	2548	0.782	ug/l	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
Data File : VX034512.D
Acq On : 15 Mar 2023 16:45
Operator : JC/MD
Sample : 01922-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-05B-69-74-031423

Quant Time: Mar 16 05:21:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

