

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
 Data File : VX034515.D
 Acq On : 15 Mar 2023 17:56
 Operator : JC/MD
 Sample : 01909-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DS-01-031023-AM

Quant Time: Mar 16 05:22:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	245963	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	419785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193687	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	167106	46.700	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.400%
35) Dibromofluoromethane	5.385	113	130971	47.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.500%
50) Toluene-d8	8.647	98	510337	49.324	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.640%
62) 4-Bromofluorobenzene	11.079	95	199262	47.679	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
Data File : VX034515.D
Acq On : 15 Mar 2023 17:56
Operator : JC/MD
Sample : 01909-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DS-01-031023-AM

Quant Time: Mar 16 05:22:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

