

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
 Data File : VX034517.D
 Acq On : 15 Mar 2023 18:43
 Operator : JC/MD
 Sample : 01880-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20230310

Quant Time: Mar 16 05:22:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	239988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	409436	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	361029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159895	45.798	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.600%
35) Dibromofluoromethane	5.385	113	127424	47.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	8.646	98	493799	48.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	184044	45.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.300%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.330	101	341187	127.898	ug/l	99
12) 1,1-Dichloroethene	2.318	96	3242	1.234	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	5821	1.733	ug/l	87
30) Chloroform	5.098	83	1591	0.297	ug/l	85
38) Carbon Tetrachloride	5.684	117	1383	0.335	ug/l	97
44) Trichloroethene	7.122	130	388938	123.240	ug/l	99
64) Tetrachloroethene	9.274	164	32516	13.408	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
Data File : VX034517.D
Acq On : 15 Mar 2023 18:43
Operator : JC/MD
Sample : 01880-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20230310

Quant Time: Mar 16 05:22:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

