

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
 Data File : VX034523.D
 Acq On : 15 Mar 2023 21:04
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 16 05:25:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	244161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	415876	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150036	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161840	45.562	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.120%
35) Dibromofluoromethane	5.385	113	128543	46.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.620%
50) Toluene-d8	8.647	98	496572	48.445	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.880%
62) 4-Bromofluorobenzene	11.079	95	165443	39.959	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.920%#
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1319	0.406	ug/l	91
44) Trichloroethene	7.123	130	3095	0.966	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
Data File : VX034523.D
Acq On : 15 Mar 2023 21:04
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 16 05:25:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

