

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
 Data File : VX034525.D
 Acq On : 15 Mar 2023 21:51
 Operator : JC/MD
 Sample : 01904-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-20230313

Quant Time: Mar 16 05:26:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	239636	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	407127	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	388133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162241	46.538	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.080%
35) Dibromofluoromethane	5.385	113	126905	47.211	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	8.647	98	487560	48.588	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	205029	50.584	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.160%
Target Compounds						
16) Acetone	2.374	43	4018	2.447	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031523\
Data File : VX034525.D
Acq On : 15 Mar 2023 21:51
Operator : JC/MD
Sample : 01904-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-20230313

Quant Time: Mar 16 05:26:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

