

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031524\
 Data File : VX040666.D
 Acq On : 15 Mar 2024 20:35
 Operator : JC/MD
 Sample : P1754-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YCKT6

Quant Time: Mar 16 05:52:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 05:02:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	147492	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	139343	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	67370	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	47420	36.677	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	73.360%		
7) Chloroethane-d5	1.666	69	44237	42.857	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	85.720%		
11) 1,1-Dichloroethene-d2	2.306	65	22180	37.267	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.540%		
21) 2-Butanone-d5	4.452	46	99724	100.515	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	100.510%		
24) Chloroform-d	5.056	84	110259	45.922	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.840%		
26) 1,2-Dichloroethane-d4	5.958	65	83561	46.615	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.220%		
32) Benzene-d6	5.976	84	188053	43.989	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.980%		
36) 1,2-Dichloropropane-d6	7.306	67	61430	46.365	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	92.720%		
41) Toluene-d8	8.647	98	167819	40.734	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	81.460%		
43) trans-1,3-Dichloroprop...	8.952	79	30871	43.778	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	87.560%		
47) 2-Hexanone-d5	9.384	63	70688	96.692	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	96.690%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	87298	44.007	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	88.020%		
66) 1,2-Dichlorobenzene-d4	12.323	152	67917	48.133	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.260%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.312	96	1830	2.036	ug/L	# 1
13) Acetone	2.380	43	50387	71.119	ug/L	97
19) 1,1-Dichloroethane	3.611	63	5903	3.292	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031524\
Data File : VX040666.D
Acq On : 15 Mar 2024 20:35
Operator : JC/MD
Sample : P1754-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YCKT6

Quant Time: Mar 16 05:52:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 05:02:38 2024
Response via : Initial Calibration

