

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008137.D
 Acq On : 16 Mar 2019 13:34
 Operator : JC/SP
 Sample : PB117440TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117440TB

Quant Time: Mar 18 01:33:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	303326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	408975	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	347483	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135739	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	
35) Dibromofluoromethane	5.50	113	125146	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	8.71	98	477401	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
62) 4-Bromofluorobenzene	11.13	95	150740	42.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.32%	
Target Compounds						
16) Acetone	2.45	43	5326	3.687	ug/l	93
20) Methylene Chloride	2.86	84	15907	6.166	ug/l	90
43) Isopropyl Acetate	6.46	43	9625	1.653	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008137.D
Acq On : 16 Mar 2019 13:34
Operator : JC/SP
Sample : PB117440TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117440TB

Quant Time: Mar 18 01:33:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

