

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008140.D
 Acq On : 16 Mar 2019 14:54
 Operator : JC/SP
 Sample : K1905-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT5124

Quant Time: Mar 18 01:42:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	424559	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	372150	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	171847	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	139476	44.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.20%	
35) Dibromofluoromethane	5.50	113	128421	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
50) Toluene-d8	8.71	98	495016	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	11.13	95	163728	44.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.26%	
Target Compounds						
16) Acetone	2.45	43	10485	7.022	ug/l	98
18) Methyl Acetate	2.78	43	6947	2.153	ug/l	92
20) Methylene Chloride	2.86	84	10054	3.770	ug/l	91
43) Isopropyl Acetate	6.46	43	9571	1.583	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008140.D
Acq On : 16 Mar 2019 14:54
Operator : JC/SP
Sample : K1905-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT5124

Quant Time: Mar 18 01:42:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

