

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031619\
 Data File : VX008143.D
 Acq On : 16 Mar 2019 16:14
 Operator : JC/SP
 Sample : K1698-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-04-031419

Quant Time: Mar 18 02:04:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	298725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	402805	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	346795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158181	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	132671	44.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.06%	
35) Dibromofluoromethane	5.50	113	122387	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
50) Toluene-d8	8.71	98	467298	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
62) 4-Bromofluorobenzene	11.13	95	150640	43.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.56%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	12137	8.531	ug/l	96
18) Methyl Acetate	2.78	43	9433	3.068	ug/l	91
20) Methylene Chloride	2.86	84	36719	14.452	ug/l	93
43) Isopropyl Acetate	6.46	43	9184	1.601	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	9962	1.157	ug/l	99
95) Naphthalene	13.83	128	184706	18.991	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031619\
Data File : VX008143.D
Acq On : 16 Mar 2019 16:14
Operator : JC/SP
Sample : K1698-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-04-031419

Quant Time: Mar 18 02:04:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

